

TOPICAL REVIEW

A survey of signal processing algorithms in brain–computer interfaces based on electrical brain signals

Ali Bashashati¹, Mehrdad Fatourech¹, Rabab K Ward^{1,2}
and Gary E Birch^{1,2,3}

¹ Department of Electrical and Computer Engineering, The University of British Columbia, 2356 Main Mall, Vancouver, V6T 1Z4, Canada

² Institute for Computing, Information and Cognitive Systems, 289-2366 Main Mall, Vancouver, BC V6T 1Z4, Canada

³ Neil Squire Society, 220-2250 Boundary Rd, Burnaby, BC, V5M 4L9, Canada

E-mail: alibs@ece.ubc.ca

Received 11 October 2006

Accepted for publication 16 February 2007

Published 27 March 2007

Online at stacks.iop.org/JNE/4/R32

Abstract

Brain–computer interfaces (BCIs) aim at providing a non-muscular channel for sending commands to the external world using the electroencephalographic activity or other electrophysiological measures of the brain function. An essential factor in the successful operation of BCI systems is the methods used to process the brain signals. In the BCI literature, however, there is no comprehensive review of the signal processing techniques used. This work presents the first such comprehensive survey of all BCI designs using electrical signal recordings published prior to January 2006. Detailed results from this survey are presented and discussed. The following key research questions are addressed: (1) what are the key signal processing components of a BCI, (2) what signal processing algorithms have been used in BCIs and (3) which signal processing techniques have received more attention?

 This article has associated online supplementary data files

1. Introduction

The ultimate purpose of a direct brain–computer interface (BCI) is to allow an individual with severe motor disabilities to have effective control over devices such as computers, speech synthesizers, assistive appliances and neural prostheses. Such an interface would increase an individual's independence, leading to an improved quality of life and reduced social costs.

A BCI system detects the presence of specific patterns in a person's ongoing brain activity that relates to the person's intention to initiate control. The BCI system translates these patterns into meaningful control commands. To detect these patterns, various signal processing algorithms are employed.

Signal processing forms an important part of a BCI design, since it is needed in extracting the meaningful information from the brain signal.

This paper summarizes the results of a comprehensive survey of different signal processing schemes that have been used in BCI systems. It specifically focuses on the following signal processing components of a BCI: the pre-processing (which we refer to as the signal enhancement), feature extraction, feature selection/dimensionality reduction, feature classification and post-processing blocks. To address all related BCI research, we include all the approaches that use standard scalp-recorded EEG as well as those that use epidural, subdural or intracortical recordings. The aims of this study are (a) to make it easy to identify the signal




Figure 1. Functional model of a BCI system (Mason and Birch 2003). Note that the control display is optional. This review focuses on the shaded components of BCI systems.

processing methods employed in different BCI systems, and consequently to identify the methods that have not yet been explored, (b) to form a historical reference for new researchers in this field and (c) to introduce a possible taxonomy of signal processing methods in brain–computer interfaces.

The organization of the paper is as follows. In section 2, the general structure of a BCI system and the current neuromechanisms⁴ in BCI systems are presented. Section 3 details the procedure we followed to conduct this study. Results, discussion and conclusion are in sections 4–6, respectively.

2. General structure of a BCI system

Figure 1 shows the functional model of a BCI system (Mason and Birch 2003). The figure depicts a generic BCI system in which a person controls a device in an operating environment (e.g., a powered wheelchair in a house) through a series of functional components. In this context, the user's brain activity is used to generate the control signals that operate the BCI system. The user monitors the state of the device to determine the result of his/her control efforts. In some systems, the user may also be presented with a control display, which displays the control signals generated by the BCI system from his/her brain activity.

The electrodes placed on the head of the user record the brain signal from the scalp, or the surface of the brain, or from the neural activity within the brain, and convert this brain activity to electrical signals. The 'artifact processor' block shown in figure 1 removes the artifacts from the electrical signal after it has been amplified. Note that many transducer designs do not include artifact processing. The 'feature generator' block transforms the resultant signals into feature values that correspond to the underlying neurological mechanism employed by the user for control. For example, if the user is to control the power of his/her mu (8–12 Hz) and beta (13–30 Hz) rhythms, the feature generator

would continually generate features relating to the power-spectral estimates of the user's mu and beta rhythms. The feature generator generally can be a concatenation of three components—the 'signal enhancement', the 'feature extraction' and the 'feature selection/dimensionality reduction' components, as shown in figure 1.

In some BCI designs, pre-processing is performed on the brain signal prior to the extraction of features so as to increase the signal-to-noise ratio of the signal. In this paper, we use the term 'signal enhancement' to refer to the pre-processing stage. A feature selection/dimensionality reduction component is sometimes added to the BCI system after the feature extraction stage. The aim of this component is to reduce the number of features and/or channels used so that very high dimensional and noisy data are excluded. Ideally, the features that are meaningful or useful in the classification stage are identified and chosen, while others (including outliers and artifacts) are omitted.

The 'feature translator' translates the features into logical (device-independent) control signals, such as a two-state discrete output. The translation algorithm uses linear classification methods (e.g., classical statistical analyses) or nonlinear ones (e.g., neural networks). According to the definition in Mason and Birch (2003), the resultant logical output states are independent of any semantic knowledge about the device or how it is controlled. As shown in figure 1, a feature translator may consist of two components: 'feature classification' and 'post-processing'. The main aim of the feature classification component is to classify the features into logical control signals. Post-processing methods such as a moving average block may be used after feature classification to reduce the number of error activations of the system. The components between the user and control interface can be treated as a single component, a BCI transducer, which functions in a manner similar to physical transducers like a dial or switch. The role of the BCI transducer is to translate the user's brain activity into logical (or device-independent) control signals.

The control interface translates the logical control signals (from the feature translator) into semantic control signals that are appropriate for the particular type of device used.

⁴ According to the Merriam-Webster Medical Dictionary, a bodily regulatory mechanism based in the structure and functioning of the nervous system is called a *neuromechanism*.

Table 1. Taxonomy for BCI transducer designs.

Item	Terms	Definition	
BCI transducer	Artifact processor	Removes artifact from the input signal	
	Feature generator	Signal enhancement	(1) Enhances signal-to-noise ratio of the brain signal (2) The output of this block is a signal with the same nature of the input (i.e. the output like the input is in the temporal domain).
		Feature extraction Feature selection/ dimensionality reduction	Generates the feature vectors Selects a subset or reduces the dimensionality of features
	Feature translator	Feature classification	Classifies the features into logical control signals
Post-processing		Increases the performance after feature classification, e.g., by blocking activations with low certainty	

Finally, the device controller translates the semantic control signals into physical control signals that are used by the device. The device controller also controls the overall behavior of the device. For more detail, refer to Mason and Birch (2003).

Table 1 provides a simplified description of the BCI transducer components.

2.1. Electrophysiological sources of control in current BCIs

In BCI systems, electrophysiological sources refer to the neurological mechanisms or processes employed by a BCI user to generate control signals. Current BCIs fall into seven main categories, based on the *neuromechanisms* and *recording technology* they use. In Wolpaw *et al* (2002) BCI systems are categorized as five major groups. These categories are sensorimotor activity, P300, VEP, SCP and activity of neural cell (ANC). In this paper, two other categories were added: ‘response to mental tasks’ and ‘multiple neuromechanisms’. BCI systems that use non-movement mental tasks to control a BCI (e.g. Anderson *et al* (1995b) and Millan *et al* (1998)) assume that different mental tasks (e.g. solving a multiplication problem, imagining a 3D object, or mental counting) lead to distinct, task-specific EEG patterns and aim to detect the patterns associated with these mental tasks from the EEG. BCI systems based on multiple neuromechanisms (e.g. Gysels *et al* (2005)) use a combination of two or more of the above-mentioned neuromechanisms in a single design of a BCI system.

Table 2 shows these categories with a short description of each. Note that although the designs that use direct cortical recordings are included as a separate group, direct cortical recording is a recording technology and not a neuromechanism. As shown in table 2, BCI designs that use sensorimotor activity as the neural source of control can be further divided into three sub-categories: those based on changes in brain rhythms (e.g. the mu and beta rhythms), those based on movement-related potentials (MRPs) and those based on other sensorimotor activity.

3. Methods

The BCI designs selected for this review include every journal and conference paper that met the following criteria:

- (1) One or more of the keywords BCI, BMI, DBI appeared in its title, abstract or keyword list.
- (2) The work described one or more BCI designs (the minimum design content that met the criteria was a *BCI transducer* as described in section 2). There were a few papers that only reported pre-processing techniques specifically designed for brain–computer interfaces that use neural cortical recordings. These papers were reported in the pre-processing techniques. Papers that presented tutorials, descriptions of electrode technology, neuroanatomy, and neurophysiology discussions that might serve as the basis for a BCI were not included.
- (3) Only papers published in English and in refereed international journals and conference proceedings were included.
- (4) Designs that use functional magnetic resonance imaging (fMRI) (Weiskopf *et al* 2003, 2004, Yoo *et al* 2004), magneto-encephalography (MEG) signals (Georgopoulos *et al* 2005), near-infra-red spectrum (NIRS), auditory evoked potentials (Hill *et al* 2004, Su Ryu *et al* 1999) and somatosensory evoked potentials (Yan *et al* 2004) were not included in this paper.
- (5) Papers were published prior to January 2006.

Although no paper meeting the five criteria explained above was omitted from the analysis, some papers may have been missed unintentionally. The current work should thus be regarded as an initial step to build a public database that can be updated and evolved with time.

In tables 3–8, we categorized the papers according to the signal processing methods used. For each of the design blocks of a BCI system shown in figure 1 (signal enhancement, feature extraction, feature selection/dimensionality reduction, feature classification, and post-processing), we created a table that reports the signal processing techniques used in that block. Since most of the BCI designs that use neural cortical recordings do not contain a feature extraction component, we generated a separate table (table 7) to report the translation schemes used in these designs.

Feature extraction methods used in BCI systems are closely related to the specific neuromechanism(s) used by a BCI. For example, the feature extraction algorithms employed in VEP-based BCIs are used to detect the visual evoked potentials in the ongoing EEG. In BCI systems that operate on slow cortical potentials (SCP), the extracted features are mostly used for the purpose of identifying this specific

Table 2. Electrophysiological activities used in BCI designs and their definitions.

Neuromechanism	Short description
Sensorimotor activity	Mu rhythms in the range of 8–12 Hz and beta rhythms in the range of 13–30 Hz both originate in the sensorimotor cortex and are displayed when a person is not engaged in processing sensorimotor inputs or in producing motor outputs (Jasper and Penfield 1949). They are mostly prominent in frontal and parietal locations (Kozelka and Pedley 1990, Kubler <i>et al</i> 2001a, Niedermeyer and Lopes da Silva 1998). A voluntary movement results in a circumscribed desynchronization in the mu and lower beta bands (Pfurtscheller and Aranibar 1977). This desynchronization is called event-related desynchronization (ERD) and begins in the contralateral rolandic region about 2 s prior to the onset of a movement and becomes bilaterally symmetrical immediately before execution of movement (Pfurtscheller and Lopes da Silva 1999). After a voluntary movement, the power in the brain rhythms increases. This phenomenon, called event-related synchronization (ERS), is dominant over the contralateral sensorimotor area and reaches a maximum around 600 ms after movement offset (Pfurtscheller and Lopes da Silva 1999). Gamma rhythm is a high-frequency rhythm in the EEG. Upon the occurrence of a movement, the amplitude of gamma rhythm increases. Gamma rhythms are usually more prominent in the primary sensory area.
Movement-related potentials (MRPs)	MRPs are low-frequency potentials that start about 1–1.5 s before a movement. They have bilateral distribution and present maximum amplitude at the vertex. Close to the movement, they become contralaterally preponderant (Babiloni <i>et al</i> 2004, Deecke and Kornhuber 1976, Hallett 1994).
Other sensorimotor activities	The sensorimotor activities that do not belong to any of the preceding categories are categorized as other sensorimotor activities. These activities are usually not restricted to a particular frequency band or scalp location and usually cover different frequency ranges. An example would be features extracted from an EEG signal filtered to frequencies below 30 Hz. Such a range covers different event-related potentials (ERPs) but no specific neuromechanism is used.
Slow cortical potentials (SCPs)	SCPs are slow, non-movement potential changes generated by the subject. They reflect changes in cortical polarization of the EEG lasting from 300 ms up to several seconds. Functionally, a SCP reflects a threshold regularization mechanism for local excitatory mobilization (Neumann <i>et al</i> 2003, Wolpaw <i>et al</i> 2002).
P300	Infrequent or particularly significant auditory, visual, or somatosensory stimuli, when interspersed with frequent or routine stimuli, typically evoke in the EEG over the parietal cortex a positive peak at about 300 ms after the stimulus is received. This peak is called P300 (Allison and Pineda 2003, Kubler <i>et al</i> 2001a).
Visual evoked potentials (VEPs)	VEPs are small changes in the ongoing brain signal. They are generated in response to a visual stimulus such as flashing lights and their properties depend on the type of the visual stimulus (Kubler <i>et al</i> 2001a). These potentials are more prominent in the occipital area. If a visual stimulus is presented repetitively at a rate of 5–6 Hz or greater, a continuous oscillatory electrical response is elicited in the visual pathways. Such a response is termed steady-state visual evoked potentials (SSVEP). The distinction between VEP and SSVEP depends on the repetition rate of the stimulation (Gao <i>et al</i> 2003b).
Response to mental tasks	BCI systems based on non-movement mental tasks assume that different mental tasks (e.g., solving a multiplication problem, imagining a 3D object, and mental counting) lead to distinct, task-specific distributions of EEG frequency patterns over the scalp (Kubler <i>et al</i> 2001a).
Activity of neural cells (ANC)	It has been shown that the firing rates of neurons in the motor cortex are increased when movements are executed in the preferred direction of neurons. Once the movements are away from the preferred direction of neurons, the firing rate is decreased (Donoghue 2002, Olson <i>et al</i> 2005).
Multiple neuromechanisms (MNs)	BCI systems based on multiple neuromechanisms use a combination of two or more of the above-mentioned neuromechanisms.

^a In Ramachandran and Hirstein (1998), references regarding the similarity between *attempted movements* and *real movements* in the ERD of mu patterns are provided. An attempted movement occurs when a subject attempts to move some part of his/her body, but because of either a disability or the experiment control, the actual movement does not happen. Similarly, Gevins *et al* (1989) have shown that imaginary (attempted) movements generate movement-related potentials (MRPs) similar to those generated by actual movements. Thus, neuromechanisms corresponding to attempted movements are grouped in the same category of real movement.

phenomenon in the brain signal. Thus in table 5, we categorize the feature extraction algorithms based on the seven neurological sources described in table 2. For example, the methods used in VEP-based BCIs are assembled under a different category from those used in SCP-based BCIs. A more detailed version of table 5 can be found in appendix B at stacks.iop.org/JNE/4/R32.

The different feature classification algorithms used in BCI systems are shown in table 6. As feature classification algorithms are also closely related to the type of the features that they classify, the feature classification

algorithms are also categorized based on the feature extraction methods. This can be found in the supplementary data at stacks.iop.org/JNE/4/R32, which also contains a detailed version of table 7, where the classification algorithms for BCIs that use cortical neural recordings are shown.

Categorizing the feature classification methods based on the feature extraction methods used does not necessarily limit the use of a specific feature classification to a specific feature extraction method. The same applies to the categorization of the feature extraction methods based on neuromechanisms used in BCI systems. The aim here is to provide as specific

information as possible about signal processing in current BCI designs and the researchers can combine any feature extraction method and/or feature classification method from different categories if necessary.

Each table includes major classes corresponding to each design block. These classes were initially determined by our team and then refined after an initial pass through the selected papers. In some cases, each major class was further divided into more specific categories. The full classification template with all the major classes and sub-classes of each design component is listed in the left column of tables 3, 4 and 8, and in the left two major columns of tables 5–7. Note that in this paper, major classes are written in **bold** type and sub-classes are represented in *bold-italic* type. For example, in table 5 or appendix B (in supplementary data), VEP-based BCI designs that use some type of power-spectral parameters of the EEG are categorized under the **VEP-spectral parameters** class, while a BCI design that is based on the movement-related potentials (MRP) and that uses the same method is categorized under the **sensorimotor activity-spectral parameters** and **sensorimotor activity-MRP-spectral parameters** classes in table 5 and appendix B, respectively. As an example from table 6, BCI designs that use linear discriminant analysis (LDA) classifiers are categorized under **LDA**. Supplementary data at stacks.iop.org/JNE/4/R32, which has a more detailed version of table 6, categorizes BCI designs that use PSD features and LDA classifier under **PSD-LDA** class.

The category for each BCI design was determined by selecting the closest sub-class in the classification template. For the papers that reported multiple designs multiple classifications were recorded. The designs were categorized based only on what was reported in each paper. No personal knowledge of an authors' related work was used in the classification.

In some cases, it was difficult to differentiate between the signal enhancement, feature selection and feature extraction design components of a brain–computer interface. Based on the definitions in table 1, the methods that satisfied the following four criteria were considered to be signal enhancement methods:

- (1) The method was implemented to improve the signal-to-noise ratio of the brain signal.
- (2) The output of the block had the same nature as the input brain signal (i.e. the output stayed in the temporal domain).
- (3) The algorithm was directly performed on the brain signal and not on the features extracted from the brain signal.
- (4) The method did not handle artifacts.

The common spatial patterns (CSP) method is an example of a method that satisfies the four above-mentioned criteria and was thus categorized as a signal enhancement method. The principle component analysis (PCA) method is another example that *sometimes* satisfies the above four criteria and was categorized as a signal-enhancement method. In the cases where PCA is applied after feature extraction to reduce the dimensionality of the extracted *features*, it is categorized as feature selection/dimensionality reduction method. Only

designs that incorporated signal enhancement algorithms other than the general band-pass filtering of the EEG, the power-line-effect rejection and the traditional normalization of the signal were reported in the signal enhancement section of this paper.

4. Results

The detailed classification results of the survey are summarized in tables 3–8 (refer to supplementary data at stacks.iop.org/JNE/4/R32 for more detailed versions of tables 5–7). As mentioned in section 3, these six tables address the signal enhancement, feature selection/dimensionality reduction, feature extraction, feature classification and post-processing methods used. The references listed for each sub-class entry represent all the papers that reported on designs related to that sub-class. As such, one can find all the designs that have specific attributes of interest. For example, if one is interested in all BCI technology designs that have used parametric modeling (and specifically extracted AR parameters of the signal) to detect the sensorimotor activity, then all the references to the relevant papers can be found in table 5 under **sensorimotor activity—parametric modeling (AR, AAR and ARX parameters)**. Alternatively, if one is looking for designs that do not have a feature extraction block but directly apply the support vector machine (SVM) classification method on the brain signal, then these papers can be easily located in the **none-SVM** class in supplementary data at stacks.iop.org/JNE/4/R32. Similarly, the designs that use the SVM classification method, regardless of the feature extraction technique used, are categorized under the **SVM** class in table 6.

To enhance the clarity of tables 3–8, the following notations are used:

- (a) BCI designs based on multiple neuromechanisms (as defined in table 2) are presented in separate categories such as **MN: sensorimotor activity + response to mental tasks**, which show that the design is based on the sensorimotor activity and the response to mental tasks neuromechanisms.
- (b) Two or more methods that are consecutively used in a design block are separated by ‘–’. As an example, **CSP–log transformation** denotes a design that first applies common spatial patterns (CSP) on the signal and then applies a logarithmic function on the resulting time-series.
- (c) Two methods that are applied simultaneously in a design component are separated by ‘+’. For example, **AR parameters + PSD parameters** corresponds to a design that uses both autoregressive (AR) and power-spectral-density (PSD) features in the feature extraction block.
- (d) **LRP + {CSP – log-transformation}** denotes designs that use the two kinds of feature extraction methods separated by ‘+’. The first method is based on the extraction of lateralized readiness potentials (LRP), and the second feature extraction method is based on consecutively applying CSP followed by a logarithm function on the signals that are grouped in ‘{ }’.
- (e) To facilitate readability, we have provided an index of terms in appendix A.

Table 3. Pre-processing (signal enhancement) methods in BCI designs.

Pre-processing method	Reference ID
Common average referencing (CAR)	(Cheng <i>et al</i> 2004, Fabiani <i>et al</i> 2004, Kubler <i>et al</i> 2005, Li <i>et al</i> 2004b, McFarland and Wolpaw 1998, McFarland <i>et al</i> 1997, 2003, Peters <i>et al</i> 2001, Ramoser <i>et al</i> 2000, Schalk <i>et al</i> 2000, Wolpaw <i>et al</i> 1997)
Surface Laplacian (SL)	(Babiloni <i>et al</i> 2000, 2001b, Cincotti <i>et al</i> 2001, 2003a, Dornhege <i>et al</i> 2004, Fabiani <i>et al</i> 2004, Gysels and Celka 2004, McFarland and Wolpaw 1998, McFarland <i>et al</i> 1997, 2003, 2005, Millan <i>et al</i> 1998, 2000a, 2002a, 2002b, 2003a, 2004a, 2004b, Millan 2004, Millan and Mourino 2003b, Muller <i>et al</i> 2003b, Peters <i>et al</i> 2001, Qin <i>et al</i> 2004a, 2004b, 2005, Qin and He 2005, Ramoser <i>et al</i> 2000, Schalk <i>et al</i> 2000, Wang <i>et al</i> 2004a, 2004b, Wolpaw and McFarland 2004)
Independent component analysis (ICA)	(Bayliss and Ballard 1999, 2000a, 2000b, Erfanian and Erfani 2004, Gao <i>et al</i> 2004, Peterson <i>et al</i> 2005, Wu <i>et al</i> 2004, Serby <i>et al</i> 2005, Xu <i>et al</i> 2004a, Wang <i>et al</i> 2004c, Li <i>et al</i> 2004a)
Common spatial patterns (CSP)	(Blanchard and Blankertz 2004, Dornhege <i>et al</i> 2003, 2004, Guger <i>et al</i> 2000b, Krauledat <i>et al</i> 2004, Muller <i>et al</i> 2003b, Pfurtscheller <i>et al</i> 2000, Pfurtscheller and Neuper 2001, Ramoser <i>et al</i> 2000, Townsend <i>et al</i> 2004, Xu <i>et al</i> 2004b)
Principal component analysis (PCA)	(Chapin <i>et al</i> 1999, Guan <i>et al</i> 2004, Hu <i>et al</i> 2004, Isaacs <i>et al</i> 2000, Lee and Choi 2002, 2003, Thulasidas <i>et al</i> 2004, Xu <i>et al</i> 2004a, Yoon <i>et al</i> 2005, Li <i>et al</i> 2004a)
Combined CSP and PCA	(Xu <i>et al</i> 2004b)
Singular value decomposition (SVD)	(Trejo <i>et al</i> 2003)
Common spatio-spatial patterns (CSSP)	(Lemm <i>et al</i> 2005)
Frequency normalization (Freq-Norm)	(Bashashati <i>et al</i> 2005, Borisoff <i>et al</i> 2004, Fatourechi <i>et al</i> 2004, 2005, Yu <i>et al</i> 2002)
Local averaging technique (LAT)	(Peters <i>et al</i> 2001)
Robust Kalman filtering	(Bayliss and Ballard 1999, 2000a)
Common spatial subspace decomposition (CSSD)	(Cheng <i>et al</i> 2004, Li <i>et al</i> 2004b, Liu <i>et al</i> 2003, Wang <i>et al</i> 2004d)
Wiener filtering	(Vidal 1977)
Sparse component analysis	(Li <i>et al</i> 2004a)
Maximum noise fraction (MNF)	(Peterson <i>et al</i> 2005)
Spike detection methods	(Obeid and Wolf 2004)
Neuron ranking methods	(Sanchez <i>et al</i> 2004)

Table 4. Feature selection/dimensionality reduction methods in BCI designs.

Feature selection/dimensionality reduction method	Reference ID
Genetic algorithm (GA)	(Flotzinger <i>et al</i> 1994, Garrett <i>et al</i> 2003, Graimann <i>et al</i> 2003a, 2004, Peterson <i>et al</i> 2005, Scherer <i>et al</i> 2004, Schroder <i>et al</i> 2003, Tavakolian <i>et al</i> 2004, Yom-Tov and Inbar 2001, 2002)
Principal component analysis (PCA)	(Anderson <i>et al</i> 1995a, Bashashati <i>et al</i> 2005, Borisoff <i>et al</i> 2004, Fatourechi <i>et al</i> 2004, 2005)
Distinctive sensitive learning vector quantization (DSLVOQ)	(Flotzinger <i>et al</i> 1994, Neuper <i>et al</i> 2005, Pfurtscheller <i>et al</i> 1996, 1997, 1998, 2000, Pfurtscheller and Neuper 2001, Pregoner and Pfurtscheller 1995, 1999)
Sequential forward feature selection (SFFS)	(Fabiani <i>et al</i> 2004, Keirn and Aunon 1990)
Grid search method	(Glassman 2005)
Relief method (Kira and Rendell 1992)	(Millan <i>et al</i> 2002a)
Recursive feature/channel elimination (RFE)	(Lal <i>et al</i> 2004, Schröder <i>et al</i> 2005)
Support vector machine (SVM)-based recursive feature elimination	(Gysels <i>et al</i> 2005)
Stepwise discriminant procedure	(Vidal 1977)
Linear discriminant analysis (LDA)	(Graimann <i>et al</i> 2003b)
Fisher discriminant analysis (dimensionality reduction)	(Wang <i>et al</i> 2004d)
Fisher discriminant-based criterion (feature selection)	(Cheng <i>et al</i> 2004, Lal <i>et al</i> 2004)
Zero-norm optimization (l0-opt)	(Lal <i>et al</i> 2004)
Orthogonal least square (OLS1) based on radial basis function (RBF)	(Xu <i>et al</i> 2004b)

5. Discussion

Several points raised in the previous section deserve further comment. Figure 2 summarizes the information in tables 3, 4 and 8, which respectively address signal

enhancement, feature selection/dimensionality reduction and post-processing algorithms in BCI designs. Specifically, this figure shows the number of BCI designs that use specific signal enhancement, feature selection/dimensionality reduction and post-processing techniques.

Table 5. Feature extraction methods in BCI designs. Refer to appendix B in supplementary data for a more detailed version of this table.

Neuro-mechanism	Feature extraction method	Reference ID
Sensorimotor activity	Spectral parameters	(Babiloni <i>et al</i> 2000, 2001a, 2001b, Blanchard and Blankertz 2004, Boostani and Moradi 2004, Cho <i>et al</i> 2004, Cincotti <i>et al</i> 2001, 2003a, 2003b, Coyle <i>et al</i> 2005, Fabiani <i>et al</i> 2004, Flotzinger <i>et al</i> 1994, Garcia <i>et al</i> 2003b, Garrett <i>et al</i> 2003, Guger <i>et al</i> 2000b, 2003a, Ivanova <i>et al</i> 1995, Kalcher <i>et al</i> 1992, 1993, Kelly <i>et al</i> 2002b, Krauledat <i>et al</i> 2004, Krausz <i>et al</i> 2003, Kubler <i>et al</i> 2005, Lal <i>et al</i> 2004, Leeb and Pfurtscheller 2004, Lemm <i>et al</i> 2005, Mahmoudi and Erfanian 2002, Mason and Birch 2000, McFarland and Wolpaw 1998, McFarland <i>et al</i> 1997, 2003, 2005, Millan <i>et al</i> 2002a, 2002b, Muller <i>et al</i> 2003c, Muller-Putz <i>et al</i> 2005b, Neuper <i>et al</i> 2003, 2005, Pfurtscheller <i>et al</i> 1993, 1994, 1996, 1997, 1998, 2000, 2005, Pfurtscheller and Neuper 2001, Pfurtscheller <i>et al</i> 2003a, 2003b, Pineda <i>et al</i> 2003, Pregenzer and Pfurtscheller 1995, 1999, Ramoser <i>et al</i> 2000, Schalk <i>et al</i> 2000, Scherer <i>et al</i> 2004, Sheikh <i>et al</i> 2003, Townsend <i>et al</i> 2004, Trejo <i>et al</i> 2003, Jia <i>et al</i> 2004, Wolpaw <i>et al</i> 1991, 1997, 2000, 2003, Wolpaw and McFarland 1994, 2004, Li <i>et al</i> 2004a)
	Parametric modeling (AR, AAR & ARX parameters)	(Burke <i>et al</i> 2002, 2005, Graitmann <i>et al</i> 2003b, Guger <i>et al</i> 1999, 2000a, 2003a, 2003b, Haselsteiner and Pfurtscheller 2000, Huggins <i>et al</i> 2003, Kelly <i>et al</i> 2002a, 2002b, Lal <i>et al</i> 2004, Neuper <i>et al</i> 1999, Obermaier <i>et al</i> 2001a, 2001b, Peters <i>et al</i> 2001, Pfurtscheller <i>et al</i> 1998, 2000, Pfurtscheller and Guger 1999, Pfurtscheller and Neuper 2001, Schloegl <i>et al</i> 1997a, 1997b, Schlogl <i>et al</i> 2003, Schröder <i>et al</i> 2005, Sykacek <i>et al</i> 2003, Yoon <i>et al</i> 2005)
	TFR method	(Bashashati <i>et al</i> 2005, Birch <i>et al</i> 2002, 2003, Borisoff <i>et al</i> 2004, Bozorgzadeh <i>et al</i> 2000, Costa and Cabral 2000, Fatourechi <i>et al</i> 2004, 2005, Garcia <i>et al</i> 2003a, 2003b, Glassman 2005, Graitmann <i>et al</i> 2003a, 2004, Huggins <i>et al</i> 2003, Lemm <i>et al</i> 2004, Lisogurski and Birch 1998, Mason and Birch 2000, Mason <i>et al</i> 2004, Pineda <i>et al</i> 2000, Qin <i>et al</i> 2004b, 2005, Qin and He 2005, Yom-Tov and Inbar 2003)
	CCTM	(Balbale <i>et al</i> 1999, Graitmann <i>et al</i> 2003b, 2004, Huggins <i>et al</i> 1999, 2003, Levine <i>et al</i> 1999, 2000)
	Signal envelope – cross-correlation	(Wang <i>et al</i> 2004a, 2004b)
	Hjorth parameters	(Boostani and Moradi 2004, Lee and Choi 2002, Obermaier <i>et al</i> 2001a, 2001c, Pfurtscheller and Neuper 2001)
	Signal complexity	(Boostani and Moradi 2004, Roberts <i>et al</i> 1999, Trejo <i>et al</i> 2003)
	Combination of different feature extraction methods	(Cheng <i>et al</i> 2004, Dornhege <i>et al</i> 2003, 2004, Krauledat <i>et al</i> 2004, Mahmoudi and Erfanian 2002, Muller <i>et al</i> 2003b, Yom-Tov and Inbar 2001, 2002)
	LRP features	(Blankertz <i>et al</i> 2002a, 2003, Krauledat <i>et al</i> 2004)
	Other	(Coyle <i>et al</i> 2004, Huggins <i>et al</i> 2003, Hung <i>et al</i> 2005, LaCourse and Wilson 2003, Li <i>et al</i> 2004b, Liu <i>et al</i> 2003, Mason and Birch 2000, Pineda <i>et al</i> 2000, Qin <i>et al</i> 2004a, 2005, Wang <i>et al</i> 2004d, Xu <i>et al</i> 2004b, Yom-Tov and Inbar 2003)
None	(Barreto <i>et al</i> 1996a, 1996b, Blankertz <i>et al</i> 2002a, Lee and Choi 2002, 2003, Mahmoudi and Erfanian 2002, Parra <i>et al</i> 2002, 2003a, Schroder <i>et al</i> 2003, Trejo <i>et al</i> 2003)	
SCP	Calculation of SCP amplitude	(Birbaumer <i>et al</i> 1999, 2000, Hinterberger <i>et al</i> 2003, 2004a, 2004b, 2005a, 2005b, Kaiser <i>et al</i> 2001, 2002, Kubler <i>et al</i> 1998, 1999, 2001b, Neumann <i>et al</i> 2003, 2004)
	TFR method	(Bostanov 2004, Hinterberger <i>et al</i> 2003)
P300	Mixed filter	(Hinterberger <i>et al</i> 2003)
	None	(Hinterberger <i>et al</i> 2003, Schroder <i>et al</i> 2003)
	Cross-correlation	(Bayliss and Ballard 1999, 2000a, 2000b, Farwell and Donchin 1988)
	Stepwise discriminant analysis	(Donchin <i>et al</i> 2000, Farwell and Donchin 1988)
	Matched filtering	(Serby <i>et al</i> 2005)
	PPM	(Jansen <i>et al</i> 2004)
	TFR method	(Bostanov 2004, Donchin <i>et al</i> 2000, Fukada <i>et al</i> 1998, Glassman 2005, Jansen <i>et al</i> 2004, Kawakami <i>et al</i> 1996)
	Peak picking	(Allison and Pineda 2003, 2005, Bayliss <i>et al</i> 2004, Farwell and Donchin 1988)
	Area calculation	(Farwell and Donchin 1988)
	Area and peak picking	(Kaper and Ritter 2004b, Xu <i>et al</i> 2004a)
Not mentioned (calculated P300 but details not mentioned)	(Bayliss 2003, Polikoff <i>et al</i> 1995)	
None	(Guan <i>et al</i> 2004, Jansen <i>et al</i> 2004, Kaper and Ritter 2004a, 2004b, Kaper <i>et al</i> 2004, Thulasidas <i>et al</i> 2004)	
VEP	Spectral parameters	(Cheng and Gao 1999, Cheng <i>et al</i> 2001, 2002, 2005, Gao <i>et al</i> 2003b, Kelly <i>et al</i> 2004, 2005a, 2005b, 2005c, Lalor <i>et al</i> 2005, Middendorf <i>et al</i> 2000, Muller-Putz <i>et al</i> 2005a, Wang <i>et al</i> 2004c, 2005a)
	Lock-in amplifier	(Calhoun and McMillan 1996, McMillan and Calhoun 1995, Muller-Putz <i>et al</i> 2005a)
	Asymmetry ratio of different band powers	(Su Ryu <i>et al</i> 1999)

Table 5. (Continued.)

Neuro-mechanism	Feature extraction method	Reference ID
Response to mental tasks ^a	Cross-correlation Amplitude between N2 and P2 peaks	(Sutter 1992) (Lee <i>et al</i> 2005)
	None	(Guan <i>et al</i> 2005, Vidal 1977)
	Spectral parameters	(Bashashati <i>et al</i> 2003, Keirn and Aunon 1990, Kostov and Polak 1997, Liu <i>et al</i> 2005, Millan <i>et al</i> 1998, Palaniappan <i>et al</i> 2002, Palaniappan 2005, Peterson <i>et al</i> 2005, Polak and Kostov 1997, 1998, Wang <i>et al</i> 2005a)
	Parametric modeling (AR & AAR parameters)	(Anderson <i>et al</i> 1995b, 1998, Garrett <i>et al</i> 2003, Huan and Palaniappan 2004, Keirn and Aunon 1990, Kostov and Polak 2000, Huan and Palaniappan 2005, Polak and Kostov 1998, 1999, Sykacek <i>et al</i> 2003)
	Signal Complexity Eigenvalues of correlation matrix	(Bashashati <i>et al</i> 2003, Tavakolian <i>et al</i> 2004) (Anderson <i>et al</i> 1998)
ANC	LPC using Burg's method	(Kostov and Polak 1997)
	None	(Anderson <i>et al</i> 1995a, Panuccio <i>et al</i> 2002)
	Cross-covariance – PCA	(Isaacs <i>et al</i> 2000)
	LBG vector quantization (VQ)	(Darmanjian <i>et al</i> 2003)
	Filtering – rectification – thresholding	(Karniel <i>et al</i> 2002, Kositsky <i>et al</i> 2003, Reger <i>et al</i> 2000a, 2000b)
MN: sensorimotor activity + response to mental tasks	Averaging TFR methods	(Laubach <i>et al</i> 2000, Otto <i>et al</i> 2003, Vetter <i>et al</i> 2003) (Laubach <i>et al</i> 2000, Musallam <i>et al</i> 2004)
	None (most of these designs model the relationship between neural firing rates and ‘position and/or velocity and/or acceleration’ of hand)	(Black <i>et al</i> 2003, Byron <i>et al</i> 2005, Carmena <i>et al</i> 2003, 2005, Chapin <i>et al</i> 1999, Gao <i>et al</i> 2002, 2003a, Hatsopoulos <i>et al</i> 2004, Hu <i>et al</i> 2004, Karniel <i>et al</i> 2002, Kemere <i>et al</i> 2004, Kennedy <i>et al</i> 2000, Kim <i>et al</i> 2005a, 2005b, Lebedev <i>et al</i> 2005, Olson <i>et al</i> 2005, Patil <i>et al</i> 2004, Rao <i>et al</i> 2005, Roushe <i>et al</i> 2003, Sanchez <i>et al</i> 2002a, 2002b, 2003, Serruya <i>et al</i> 2002, 2003, Taylor <i>et al</i> 2002, 2003, Wessberg <i>et al</i> 2000, Wu <i>et al</i> 2002a, 2002b)
	Spectral parameters	(Gysels and Celka 2004, Gysels <i>et al</i> 2005, Millan <i>et al</i> 2000a, 2000b, 2002b, 2003a, 2004a, Millan 2004, Millan and Mourino 2003b, Obermaier <i>et al</i> 2001d, Varsta <i>et al</i> 2000)
	Parametric modeling	(Curran <i>et al</i> 2004, Penny <i>et al</i> 2000, Roberts and Penny 2003, Sykacek <i>et al</i> 2004, Varsta <i>et al</i> 2000)
	TFR method	(Garcia and Ebrahimi 2002, Garcia <i>et al</i> 2002, 2003c, Molina <i>et al</i> 2003, Varsta <i>et al</i> 2000)
MN: SCP + other brain rhythms	Combination of different features	(Erfanian and Erfani 2004)
	PLV	(Gysels and Celka 2004, Gysels <i>et al</i> 2005)
	Mean spectral coherence	(Gysels and Celka 2004)
	None	(Mourino <i>et al</i> 2002, Rezek <i>et al</i> 2003)
	SCP calculation + power spectral parameters	(Hinterberger and Baier 2005, Mensh <i>et al</i> 2004)

^a Designs that differentiate between relaxed state and movement tasks are considered in ‘sensorimotor activity + response to mental tasks’ category.

In the remainder of this section we highlight the top three or four methods that have been used in the signal processing blocks of BCI systems (as introduced in section 2).

Of the 96 BCI designs that employ signal enhancement techniques before extracting the features from the signal, 32% use surface Laplacian (SL), 22% use either principal component analysis (PCA) or independent component analysis (ICA), 14% use common spatial patterns (CSP) and 11% use common average referencing (CAR) techniques. Thirty-eight of the reported BCI designs employ feature selection/dimensionality reduction algorithms; 26% of these 38 designs use genetic algorithms (GA), 24% use distinctive

sensitive learning vector quantization (DSLQV), and 13% use PCA.

Of the 30 BCI designs that use post-processing algorithms to reduce the amount of error in the output of the BCI system, 57% use averaging techniques and consider rejecting activations that have low certainty, 27% consider using the debounce block (or refractory period) to deactivate the output for a short period of time when a false activation is detected, and 16% use event-related negativity (ERN) signals to detect error activations.

Figure 3 summarizes the results presented in table 5, and shows the number of BCI designs that are based on

Table 6. Feature classification methods in BCI designs that use EEG and ECoG recording technology.

Feature classification method	Reference ID
Neural networks (NN)	(Anderson <i>et al</i> 1995a, 1995b, 1998, Costa and Cabral 2000, Erfanian and Erfani 2004, Fukada <i>et al</i> 1998, Garrett <i>et al</i> 2003, Haselsteiner and Pfurtscheller 2000, Huan and Palaniappan 2004, Hung <i>et al</i> 2005, Ivanova <i>et al</i> 1995, Mahmoudi and Erfanian 2002, Huan and Palaniappan 2005, Palaniappan 2005, Su Ryu <i>et al</i> 1999, Tavakolian <i>et al</i> 2004)
<i>MLP</i>	
<i>Committee of MLP NN</i>	(Millan <i>et al</i> 2000b, 2002b, Varsta <i>et al</i> 2000)
<i>FIR- MLP NN</i>	(Haselsteiner and Pfurtscheller 2000)
<i>Committee of Platt's RAN algorithm (Platt 1991)</i>	(Millan <i>et al</i> 1998)
<i>Committee of NNs trained with Adaboost</i>	(Boostani and Moradi 2004)
<i>Committee of single perceptrons with no hidden layers</i>	(Peters <i>et al</i> 2001)
<i>TBNN</i>	(Ivanova <i>et al</i> 1995)
<i>TDNN</i>	(Barreto <i>et al</i> 1996a, 1996b)
<i>LVQ</i>	(Flotzinger <i>et al</i> 1994, Ivanova <i>et al</i> 1995, Kalcher <i>et al</i> 1992, 1993, Pfurtscheller <i>et al</i> 1993, 1994, 1996, 1997, 1998, 2000, Pfurtscheller and Neuper 2001, Pregezer and Pfurtscheller 1999)
<i>kMeans – LVQ</i>	(Bashashati <i>et al</i> 2005, Birch <i>et al</i> 2002, 2003, Borisoff <i>et al</i> 2004, Bozorgzadeh <i>et al</i> 2000, Fatourechhi <i>et al</i> 2004, 2005, Lisogurski and Birch 1998, Mason and Birch 2000, Mason <i>et al</i> 2004, Yom-Tov and Inbar 2003)
<i>fART – LVQ</i>	(Borisoff <i>et al</i> 2004)
<i>DSLVO</i>	(Muller-Putz <i>et al</i> 2005a, Neuper <i>et al</i> 2005, Pregezer and Pfurtscheller 1995, 1999)
<i>Growing hierarchical SOM</i>	(Liu <i>et al</i> 2005)
<i>ALN</i>	(Kostov and Polak 1997, 1998, 1999, 2000)
<i>ANN</i>	(Cincotti <i>et al</i> 2003b)
<i>Custom designed local NN</i>	(Mourino <i>et al</i> 2002, Millan <i>et al</i> 2000a, 2002b, Millan and Mourino 2003b)
<i>Fuzzy ARTMAP</i>	(Palaniappan <i>et al</i> 2002)
<i>Single layer NN</i>	(Garcia <i>et al</i> 2002)
<i>RBF-NN</i>	(Hung <i>et al</i> 2005)
<i>Static neural classifier (Adaline)</i>	(Barreto <i>et al</i> 1996a, 1996b)
<i>Gamma NN</i>	(Barreto <i>et al</i> 1996a, 1996b)
(R) LDA^a	(Boostani and Moradi 2004, Bostanov 2004, Burke <i>et al</i> 2002, 2005, Coyle <i>et al</i> 2004, 2005, Dornhege <i>et al</i> 2003, 2004, Fabiani <i>et al</i> 2004, Fukada <i>et al</i> 1998, Garcia <i>et al</i> 2003b, Garrett <i>et al</i> 2003, Guger <i>et al</i> 2000b, 2003a, 2003b, Hinterberger <i>et al</i> 2003, Huan and Palaniappan 2004, Huggins <i>et al</i> 2003, Kelly <i>et al</i> 2002a, 2002b, 2004, 2005a, 2005c, Krauledat <i>et al</i> 2004, Krausz <i>et al</i> 2003, Lalor <i>et al</i> 2005, Leeb and Pfurtscheller 2004, Lemm <i>et al</i> 2005, Mensh <i>et al</i> 2004, Muller <i>et al</i> 2003b, 2003c, Muller-Putz <i>et al</i> 2005a, 2005b, Neuper <i>et al</i> 1999, 2003, Obermaier <i>et al</i> 2001b, Pfurtscheller <i>et al</i> 1998, 2000, 2003b, Pfurtscheller and Guger 1999, Pfurtscheller and Neuper 2001, Schloegl <i>et al</i> 1997a, 1997b, Townsend <i>et al</i> 2004, Jia <i>et al</i> 2004)
(R) FLD	(Babiloni <i>et al</i> 2001b, Blanchard and Blankertz 2004, Blankertz <i>et al</i> 2002a, 2003, Cincotti <i>et al</i> 2001, 2003a, Guger <i>et al</i> 1999, 2000a, Hung <i>et al</i> 2005, Obermaier <i>et al</i> 2001a, Pfurtscheller <i>et al</i> 2003a, Scherer <i>et al</i> 2004, Li <i>et al</i> 2004a)
Sparse FLD	(Blankertz <i>et al</i> 2002a)
MD-based classifier	(Babiloni <i>et al</i> 2001a, Cincotti <i>et al</i> 2003a, 2003b, Garcia and Ebrahimi 2002, Molina <i>et al</i> 2003)
Nonlinear discriminant function	(Fabiani <i>et al</i> 2004)
Bayes quadratic classifier	(Keirn and Aunon 1990)
Bayesian classifier (linear classifier)	(Curran <i>et al</i> 2004, Lemm <i>et al</i> 2004, Penny <i>et al</i> 2000, Roberts and Penny 2003)
Linear Bayesian decision rule	(Vidal 1977)
Linear classifier based on time-warping	(Mason and Birch 2000)
Logistic regression	(Parra <i>et al</i> 2002, 2003a)
Linear classifier (no details)	(Ramoser <i>et al</i> 2000)
Single layer Perceptron model (a linear classifier)	(Li <i>et al</i> 2004b, Wang <i>et al</i> 2004d)
Two-dimensional linear classifier trained by a non-enumerative search procedure	(Cheng <i>et al</i> 2004)
ZDA	(Hinterberger <i>et al</i> 2003)
LDS	(Lee and Choi 2002)
Gaussian classifier	(Millan 2004, Millan <i>et al</i> 2004a, 2004b, 2003a)
SSP	(Babiloni <i>et al</i> 2000, 2001a, 2001b, Cincotti <i>et al</i> 2001, 2003a, Millan <i>et al</i> 2000b, 2002b)
SOM-based SSP	(Millan <i>et al</i> 2000b, 2002b)

Table 6. (Continued.)

Feature classification method	Reference ID
HMM-based techniques	(Rezek <i>et al</i> 2003)
<i>CHMM</i>	(Panuccio <i>et al</i> 2002)
<i>AR HMM</i>	(Lee and Choi 2002, 2003)
<i>HMM + SVM</i>	(Cincotti <i>et al</i> 2003b, Lee and Choi 2003, Liu <i>et al</i> 2003, Obermaier <i>et al</i> 2001a, 2001c, 2001d, Pfurtscheller and Neuper 2001, Sykacek <i>et al</i> 2003)
<i>HMM</i>	(Blankertz <i>et al</i> 2002a, Guan <i>et al</i> 2004, Garcia <i>et al</i> 2003a, 2003b, 2003c, Garrett <i>et al</i> 2003, Glassman 2005, Gysels and Celka 2004, Gysels <i>et al</i> 2005, Hung <i>et al</i> 2005, Guan <i>et al</i> 2005, Kaper and Ritter 2004a, 2004b, Kaper <i>et al</i> 2004, Lal <i>et al</i> 2004, Peterson <i>et al</i> 2005, Schroder <i>et al</i> 2003, Schröder <i>et al</i> 2005, Thulasidas <i>et al</i> 2004, Trejo <i>et al</i> 2003, Xu <i>et al</i> 2004b, Yom-Tov and Inbar 2001, 2002, 2003, Yoon <i>et al</i> 2005)
SVM	(Ivanova <i>et al</i> 1995)
NID3	(Ivanova <i>et al</i> 1995)
CN2	(Ivanova <i>et al</i> 1995)
C4.5	(Ivanova <i>et al</i> 1995, Millan <i>et al</i> 2002a)
k-NN	(Blankertz <i>et al</i> 2002a, Pineda <i>et al</i> 2000, Pregonzer and Pfurtscheller 1999)
Threshold detector	(Allison and Pineda 2003, Balbale <i>et al</i> 1999, Bayliss and Ballard 1999, 2000a, 2000b, Bayliss 2003, Bayliss <i>et al</i> 2004, Calhoun and McMillan 1996, Cheng and Gao 1999, Cheng <i>et al</i> 2001, 2002, 2005, Donchin <i>et al</i> 2000, Farwell and Donchin 1988, Gao <i>et al</i> 2003b, Graimann <i>et al</i> 2003a, 2003b, 2004, Hinterberger <i>et al</i> 2003, Huggins <i>et al</i> 1999, 2003, Jansen <i>et al</i> 2004, Kawakami <i>et al</i> 1996, Kelly <i>et al</i> 2005b, Kostov and Polak 1997, Lee <i>et al</i> 2005, Levine <i>et al</i> 1999, Levine <i>et al</i> 2000, McMillan and Calhoun 1995, Middendorf <i>et al</i> 2000, Pfurtscheller <i>et al</i> 2005, Pineda <i>et al</i> 2003, Polak and Kostov 1997, Polikoff <i>et al</i> 1995, Qin <i>et al</i> 2004a, 2004b, Qin and He 2005, Roberts <i>et al</i> 1999, Serby <i>et al</i> 2005, Sutter 1992, Wang <i>et al</i> 2004a, 2004b, Xu <i>et al</i> 2004a, Yom-Tov and Inbar 2003)
Linear combination – threshold detector	(Townsend <i>et al</i> 2004)
Continuous feedback + threshold detector	(Birbaumer <i>et al</i> 1999, 2000, Hinterberger <i>et al</i> 2003, 2004a, 2004b, 2005a, 2005b, Kaiser <i>et al</i> 2001, 2002, Kubler <i>et al</i> 1998, 1999, 2001b, Neumann <i>et al</i> 2003, 2004)
Linear combination – continuous feedback	(Fabiani <i>et al</i> 2004, Krausz <i>et al</i> 2003, Kubler <i>et al</i> 2005, McFarland and Wolpaw 1998, McFarland <i>et al</i> 1997, 2003, 2005, Schalk <i>et al</i> 2000, Sheikh <i>et al</i> 2003, Wolpaw and McFarland 1994, 2004, Wolpaw <i>et al</i> 1997, 2000, 2003)
Continuous feedback	(Bashashati <i>et al</i> 2003, Cho <i>et al</i> 2004, LaCourse and Wilson 2003, Middendorf <i>et al</i> 2000, Trejo <i>et al</i> 2003, Wolpaw <i>et al</i> 1991)
Continuous feedback using MD	(Schlogl <i>et al</i> 2003)
Continuous audio feedback	(Hinterberger and Baier 2005)
Variational Kalman filter	(Sykacek <i>et al</i> 2004)
Static classifier that is inferred with sequential variational inference	(Curran <i>et al</i> 2004, Sykacek <i>et al</i> 2004)
Random forest algorithm	(Neuper <i>et al</i> 1999)

^a Regularization may be applied before LDA classification scheme.

sensorimotor activity, SCP, VEP, P300, activity of neural cells, ‘response to mental tasks’ and multiple neuromechanisms and use different feature extraction techniques.

Based on the results of figure 3, 41% of the BCIs that are based on the sensorimotor activity use power-spectral-density features, 16% rely on parametric modeling of the data, 13% use time–frequency representation (TFR) methods and 6% do not employ any feature extraction methods. 74% of the SCP-based BCI designs calculate SCP signals using low-pass filtering methods, and 64% of the VEP-based BCIs use power-spectral features at specific frequencies. 26% of the BCIs based on P300 calculate the peaks of the signal in a specific time window to detect the P300 component of the EEG; 22% use TFR-based methods, 22% use no feature extraction method, and 15% use cross-correlation with a specific template. 41% of the BCI

designs that use mental tasks to control a BCI use power-spectral features and 37% use parametric modeling of the input signal. As most of the BCI designs that are based on neural cortical recordings mainly try to model the direct relationship between the neural cortical recordings and movements, they do not use a feature-extraction algorithm. 45% of the BCI designs that are based on multiple neuromechanisms rely on power-spectral features, 17% use parametric modeling, and 17% use time–frequency representation (TFR) methods.

Summarizing tables 6 and 7, the number of BCI designs that use different feature classification algorithms are shown in figure 4. About 75% of the BCI designs use classification schemes that are *not* based on neural networks (NN). These are composed of those methods that

Table 7. Feature classification methods in BCI designs that are based on neural cortical recordings.

Feature classification	Reference ID
Neural networks	(Sanchez <i>et al</i> 2002a, 2002b, 2003)
<i>Recurrent MLP neural network (RNN)</i>	
<i>MLP</i>	(Kim <i>et al</i> 2005b)
<i>Feed-forward ANN</i>	(Patil <i>et al</i> 2004)
<i>ANN recurrent dynamic back-propagation ANN model</i>	(Chapin <i>et al</i> 1999)
<i>LVQ</i>	(Hatsopoulos <i>et al</i> 2004, Wessberg <i>et al</i> 2000)
<i>Other</i>	(Laubach <i>et al</i> 2000)
Support vector machine regression (SVR) model	(Kim <i>et al</i> 2005b)
Cosine tuning model (a linear model)	(Black <i>et al</i> 2003, Kemere <i>et al</i> 2004, Taylor <i>et al</i> 2002, 2003)
Linear Gaussian models (LGM) implemented by Kalman filter	(Black <i>et al</i> 2003, Gao <i>et al</i> 2003a, Patil <i>et al</i> 2004, Sanchez <i>et al</i> 2002a, Wu <i>et al</i> 2002a, 2002b)
Generalized linear models (GLA)	(Black <i>et al</i> 2003, Gao <i>et al</i> 2003a)
Generalized additive models (GAM)	(Black <i>et al</i> 2003, Gao <i>et al</i> 2003a)
Weighted linear combination of neuronal activity (Wiener filter: a linear model)	(Carmena <i>et al</i> 2005, Hatsopoulos <i>et al</i> 2004, Kim <i>et al</i> 2005a, 2005b, Lebedev <i>et al</i> 2005, Patil <i>et al</i> 2004, Sanchez <i>et al</i> 2002b, Serruya <i>et al</i> 2003, 2002)
Gamma filter (a linear model)	(Sanchez <i>et al</i> 2002b)
Mixture of multiple models based on NMF (non-negative matrix factorization)	(Kim <i>et al</i> 2005a)
Echo state networks (ESN) — optimal sparse linear mapping	(Rao <i>et al</i> 2005)
Linear model (no details mentioned)	(Carmena <i>et al</i> 2003, Wessberg <i>et al</i> 2000)
Threshold detector	(Otto <i>et al</i> 2003, Roushe <i>et al</i> 2003, Vetter <i>et al</i> 2003)
SVM	(Byron <i>et al</i> 2005, Hu <i>et al</i> 2004, Olson <i>et al</i> 2005)
Bayesian classifier	(Gao <i>et al</i> 2002, Hu <i>et al</i> 2004, Musallam <i>et al</i> 2004)
Maximum likelihood-based model	(Hatsopoulos <i>et al</i> 2004, Kemere <i>et al</i> 2004, Serruya <i>et al</i> 2003)
LPF (continuous signal)	(Karniel <i>et al</i> 2002, Kositsky <i>et al</i> 2003, Reger <i>et al</i> 2000a, 2000b)
Direct translation of firing rate to cursor movement (continuous signal)	(Kennedy <i>et al</i> 2000)
k-NN	(Isaacs <i>et al</i> 2000)
HMM	(Darmanjian <i>et al</i> 2003)

Table 8. Post-processing methods in BCI designs.

Post processing	Reference ID
ERN (event-related negativity)-based error correction	(Bayliss <i>et al</i> 2004, Blankertz <i>et al</i> 2002b, 2003, Parra <i>et al</i> 2003b, Schalk <i>et al</i> 2000)
Successive averaging and/or rejection option for ‘moderated’ posterior probabilities (choice of ‘unknown’ output state) (SA-UK)	(Anderson <i>et al</i> 1995a, Bashashati <i>et al</i> 2005, Birch <i>et al</i> 2002, Borisoff <i>et al</i> 2004, Fatourehchi <i>et al</i> 2004, 2005, Gysels and Celka 2004, Millan <i>et al</i> 1998, 2004b, Millan 2004, Millan and Mourino 2003b, 2004b, Muller-Putz <i>et al</i> 2005b, Penny <i>et al</i> 2000, Roberts and Penny 2003, Townsend <i>et al</i> 2004, Vidal 1977, Millan <i>et al</i> 2003a)
Debounce (considering refractory period)	(Bashashati <i>et al</i> 2005, Borisoff <i>et al</i> 2004, Fatourehchi <i>et al</i> 2004, 2005, Muller-Putz <i>et al</i> 2005b, Obeid and Wolf 2004, Pfurtscheller <i>et al</i> 2005, Townsend <i>et al</i> 2004)

use threshold detectors as the feature classifier or as part of the feature classification scheme (27%), linear discriminant (either LDA or FLD) classifiers (26%), those that show continuous feedback of the extracted features (16%), and those that use support vector machines (SVM) (11%). 27% of the neural-network-based classifiers are based on the multi-layer perceptrons (MLP) neural network and 39% are based on learning-vector-quantization (LVQ) classification scheme.

During our analysis of the literature, a number of salient points about the signal processing methods emerged. We think that some of these points are worth sharing with the BCI research community. In the following sections we summarize some of them. Note that these observations are based on comments made by the researchers in their published papers and are also based on the trends in the literature; they do not cover all the methods reported in the literature.




Figure 2. Signal enhancement, feature selection/dimensionality reduction and post-processing methods in BCI designs.

5.1. Signal enhancement

Signal enhancement (pre-processing) algorithms have been used for brain–computer interfaces that are based on the EEG and the activity of the neural cells (ANC), but no signal enhancement algorithms have been applied on electrocorticogram (ECoG)-based brain–computer interfaces. Given the huge difference in the characteristics of EEG and ANC, the signal enhancement algorithms used in EEG-based and ANC-based BCIs have very little overlap; only PCA has been used in both groups. While ANC-based BCIs mostly aim at spike detection, ranking, and sorting neuron activities, EEG-based ones mostly transform or select EEG channels that yield better performance. Overall, the use of a pre-processing stage before feature extraction (if applied) has been proven to be useful. The choice of a suitable pre-processing technique however is dependent on several factors such as the recording technology, number of electrodes, and neuromechanism of the BCI. Next, we discuss some of the techniques used in signal enhancement in EEG-based BCI systems. Specifically, a discussion of spatial filtering including referencing methods and common spatial patterns (CSP) is presented. These methods are among the most used techniques that have become increasingly popular in BCI studies.

5.1.1. Referencing methods. Referencing methods are considered as spatial filters. The proper selection of a spatial filter for any BCI is determined by the location and extent of the control signal (e.g. the mu rhythm) and of the various sources of EEG or non-EEG noise. The latter two are not completely defined and presumably vary considerably between different studies and both across and within individuals (McFarland *et al* 1997).

For BCIs that use the mu and beta rhythms, the common average referencing (CAR) and Laplacian methods are superior to the ear reference method. This may be because these methods use high-pass spatial filters and enhance the focal activity from the local sources (e.g. the mu and the beta rhythms) and reduce the widely distributed activity, including that resulting from distant sources (e.g. EMG, eye movements and blinks, visual alpha rhythm). Comparing the two variations of the Laplacian filtering methods (the large Laplacian and the small Laplacian), it is shown that the large Laplacian method is superior to the small Laplacian method in BCI systems that use the mu rhythm (McFarland *et al* 1997).

Although an accurate Laplacian estimate from raw potentials requires many electrodes, one study showed that the recognition rates were increased by using a small number of electrodes only (Cincotti *et al* 2003a). In this case, the linear combination of channels implementing the Laplacian estimation was likely to have caused a favorable transformation of the signals to recognize different patterns in the ongoing EEG.

5.1.2. Common spatial patterns (CSP). CSP is a signal enhancement method that detects patterns in the EEG by incorporating the spatial information of the EEG signal. Some of its features and limitations include the following.

An advantage of the CSP method is that it does not require the *a priori* selection of subject-specific frequency bands. Knowledge of these bands, however, is necessary for the band-power and frequency-estimation methods (Guger *et al* 2000b). One disadvantage of the CSP method is that it requires the use of many electrodes. However, the inconvenience of applying more electrodes can be rationalized by improved performance (Guger *et al* 2000b, Pfurtscheller *et al* 2000). The major problem in the application of CSP is its sensitivity to artifacts in the EEG. Since the covariance matrices are used as the basis for calculating the spatial filters, and are estimated with a comparatively small number of examples, a single trial contaminated with artifacts can unfortunately cause extreme changes to the filters (Guger *et al* 2000b, Ramoser *et al* 2000). Since the CSP method detects spatial patterns in the EEG, any change in the electrode positions may render the improvements in the classification accuracy gained by this method useless. Therefore, this method requires almost identical electrode positions for all trials and sessions which may be difficult to accomplish (Ramoser *et al* 2000).

5.2. Feature extraction

In this section we discuss some of the feature extraction techniques that have received more attention in BCI systems. Specifically, time and/or frequency representation methods, parametric modeling and specific techniques of modeling neural cortical recordings are discussed.




Figure 3. Feature extraction methods in BCI designs based on sensorimotor activity, VEP, P300, SCP, response to mental tasks, activity of neural cells, and multiple neuromechanisms.

5.2.1. Time and/or frequency methods. A signal, as a function of time, may be considered as a representation with perfect temporal resolution. The magnitude of the Fourier transform (FT) of the signal may be considered as a representation with perfect spectral resolution but with no temporal information. Frequency-based features have been widely used in signal processing because of their ease of application, computational speed and direct interpretation of the results. Specifically, about one-third of BCI designs have used power-spectral features. Due to the non-stationary nature of the EEG signals, these features do not provide any time domain information. Thus, mixed time–frequency representations (TFRs) that map a one-dimensional signal into a two-dimensional function of time and frequency are used to analyze the time-varying spectral content of the signals. It has been shown that TFR methods may yield performance improvements comparing to the traditional FT-based methods (e.g. Qin *et al* (2005) and Bostanov (2004)). Most of the designs that employ TFR methods use wavelet-based feature extraction algorithms. The choice of the particular wavelet used is a crucial factor in gaining useful information from wavelet analysis. Prior knowledge of the physiological activity in the brain can be useful in determining the appropriate wavelet function.

Correlative TFR (CTFR) is another time–frequency representation method that, besides the spectral information,

provides information about the time–frequency interactions between the components of the input signal. Thus, with the CTFR the EEG data samples are not independently analyzed (as in the Fourier transform case) but their relationship is also taken into account. One drawback of the CTFR resides in its relative high sensitivity to noise. Consequently, the most important values of the CTFR in terms of classification must be selected (Garcia *et al* 2003a, 2003b).

5.2.2. Parametric modeling. Parametric approaches assume the time series under analysis to be the output of a given linear mathematical model. They require an *a priori* choice of the structure and order of the signal generation mechanism model (Weitkunat 1991). The optimum model order is best estimated not only by maximizing the fitness but also by limiting the model's complexity. For noisy signals, if the model's order is too high, spurious peaks in the spectra will result. On the other hand, if the order is too low, smooth spectra are obtained (Kelly *et al* 2002a, Polak and Kostov 1998, Weitkunat 1991).

For short EEG segments, parametric modeling results in better frequency resolution and a good spectral estimate. Note that parametric modeling may yield poor estimates if the length of the EEG segments processed is too short (Birch 1988). For such modeling, there is no need for *a priori* information about potential frequency bands, and there is no need to window the data in order to decrease the spectral leakage. Also, the




Figure 4. Feature classification methods in BCI designs. ‘In/out modeling’ refers to those ANC-based BCIs that directly map the input neural recordings to the output without using a feature extraction technique.

frequency resolution does not depend on the number of data points (Guger *et al* 2003a, Polak and Kostov 1998, Weitkunat 1991). Estimating these parameters, however, is very sensitive to artifacts (Birch 1988, Guger *et al* 2003a).

Special attention should be paid to the choice of the sampling rate in parametric modeling (Weitkunat 1991), since severely oversampled signals tend to show only very small amplitude differences between successive samples. Hence, low-order models produce small prediction errors, giving the false illusion that an adequate model has been obtained. The sampling rates dictated by the Nyquist criterion are recommended.

5.2.3. Modeling the neural firing rates. Extraction algorithms for motor control operate on spike trains, recorded from a population of cortical units, mostly with the purpose of predicting arm trajectories. Several extraction methods such as linear filtering methods and neural networks have been used to determine arm movement trajectories from neural firing rates. We summarize below a few important issues in modeling the neural firing rates. A more detailed critical discussion of

extraction algorithms for cortical control of arm prosthetics can be found in Schwartz *et al* (2001).

One limitation of linear filter methods is that they rely on an *a priori* model of movement-related neuronal responses. Artificial neural network (ANN) solutions can optimize each cell’s contribution to the population prediction (Schwartz *et al* 2001).

Several of the algorithms used are based on the position of the moving limb. In the primary motor cortex at least, this parameter is more poorly represented than the velocity during movement. With most algorithms, the different sources of variability need to be specified explicitly because some sort of optimal function is being modeled to the cell response (Schwartz *et al* 2001).

The performance of the modeling techniques is constrained by their training sets and may be limited, both in terms of extrapolation beyond and interpolation within the training set when new data are applied. The success of the linear filters is due to the underlying linearity of the relationship between firing rate and movement direction. These filters are limited by the conditions used to fit their coefficients and may suffer from the same training constraints as ANNs (Schwartz *et al* 2001).

5.3. Feature selection/dimensionality reduction

Feature selection algorithms are used in BCI designs to find the most informative features for classification. This is especially useful for BCI designs with high dimensional input data, as it reduces the dimension of the feature space. Since feature selection block reduces the complexity of the classification problem, higher classification accuracies might be achieved. The experiments carried out in Flotzinger *et al* (1994) and Pregoner and Pfurtscheller (1999) show that when feature selection is used, the classification accuracy is better than when all the features are used.

Principal component analysis (PCA) and genetic algorithms (GA) are among the mostly used feature selection and/or dimensionality reduction methods in BCIs. PCA has also been widely used in pre-processing stage of BCI designs. PCA is a linear transformation that can be used for dimensionality reduction in a dataset while retaining those characteristics of the dataset that contribute most to its variance, by keeping lower-order principal components and ignoring higher-order ones. Such low-order components often contain the ‘most important’ aspects of the data. PCA has the distinction of being the optimal linear transformation for keeping the subspace that has largest variance. PCA only finds linear subspaces, works best if the individual components have Gaussian distributions, and is not optimized for class separability. One other possible application area of PCA is in classification stage, in which, PCA is applied for weighting input features. While a standard neural network, such as the multi-layer perceptrons (MLP), can do the necessary classification itself, in some cases doing a PCA in parallel and weighting input features can give better results as it simplifies the training of the rest of the system.

Unlike PCA, GAs are heuristic search techniques in the problem space. GAs typically maintain a constant-sized population of individuals which represent samples of the space to be searched. Each individual is evaluated on the basis of its overall fitness with respect to the given application domain. New individuals (samples of the search space) are produced by selecting high performing individuals to produce ‘offspring’ which retain many of the features of their ‘parents’. This eventually leads to a population that has improved fitness with respect to the given goal. Genetic algorithms have demonstrated substantial improvement over a variety of random and local search methods (De Jong 1975). This is accomplished by their ability to exploit accumulating information about an initially unknown search space in order to bias subsequent search into promising subspaces. Since GAs are basically a domain-independent search technique, they are ideal for applications where domain knowledge and theory is difficult or impossible to provide (De Jong 1975). An important step in developing a GA-based search is defining a suitable fitness function. An ideal fitness function correlates closely with the algorithm’s goal, and yet may be computed quickly. Speed of execution is very important, as a typical genetic algorithm must be iterated many, many times in order to produce a usable result for a

non-trivial problem. Definition of the fitness function is not straightforward in many cases and often is performed iteratively if the fittest solutions produced by a GA are not what is desired.

5.4. Feature classification

Linear classifiers are generally more robust than nonlinear ones. This is because linear classifiers have fewer free parameters to tune, and are thus less prone to over-fitting (Muller *et al* 2003a). In the presence of strong noise and outliers, even linear systems can fail. One way of overcoming this problem is to use regularization. Regularization helps limit (a) the influence of outliers and strong noise, (b) the complexity of the classifier and (c) the raggedness of the decision surface (Muller *et al* 2003a).

It is always desirable to avoid reliance on nonlinear classification methods, if possible, because these methods often involve a number of parameters whose values must be chosen appropriately. However, when there are large amounts of data and limited knowledge of the data, nonlinear methods are better suited in finding the potentially more complex structure in the data. In particular, when the source of the data to be classified is not well understood, using methods that are good at finding nonlinear transformation of the data is suggested. In these cases, kernel-based and neural-networks-based methods can be used to determine the transformations. Kernel-based classifiers are classification methods that apply a linear classification in some appropriate (kernel) feature space. Thus, all the beneficial properties of linear classification are maintained, but at the same time, the overall classification is nonlinear. Examples of such kernel-based classification methods are support vector machines (SVMs) and kernel Fisher discriminant (KFD) (Muller *et al* 2003a). For a more detailed critical discussion regarding linear and nonlinear classifiers in brain-computer interfaces, refer to Muller *et al* (2003a).

Some BCI designs have used classification algorithms such as FIR-MLP and TBNN that utilize temporal information of the input data (Haselsteiner and Pfurtscheller 2000, Ivanova *et al* 1995). The motivation for using such classifiers is that the patterns to be recognized are not static data but time series. Thus, the temporal information of the input data can be used to improve the classification results (Haselsteiner and Pfurtscheller 2000). Utilizing the temporal information of features is not necessarily performed directly in the classification stage, and can be done with a static classifier like MLP and a mapping of the temporal input data to static data. However, using classifiers such as FIR-MLP and TBNN that directly utilize temporal information may yield better performances as they are much better suited for exploiting temporal information contained in the time series to be classified. Regardless of the method that is used for exploiting temporal information, these approaches are preferred over static classification as they may increase the performance of BCI systems.

Using a group (committee) of classifiers rather than using a single classifier might also yield to better performances of

BCI systems. Only a few BCI designs have employed such an approach in classifying features and achieved performance improvements (Millan *et al* 2000b, 2002b, Peters *et al* 2001, Varsta *et al* 2000). The classification accuracy of the committee depends on how much unique information each committee member contributes to classification. A committee of classifiers usually yields better classification accuracy than any individual classifier could provide, and can be used to combine information from several channels, i.e., from different spatial regions (Peters *et al* 2001).

As the number of epochs available for evaluating a BCI system is small, using a technique that reduces the bias of the estimated performance on a specific dataset is highly recommended. This is especially important when different architectures of a certain design are being compared. K -fold cross-validation and statistical significance tests are especially useful for these cases (e.g. refer to Anderson *et al* (1998), Kelly *et al* (2002b), Lalor *et al* (2005), Obermaier *et al* (2001d) and Peterson *et al* (2005)). K -fold cross-validation can be used simply to estimate the generalization error of a given model, or it can be used for model selection by choosing one of several models that has the smallest estimated generalization error but it is not suitable for online evaluations. A value of 5 to 10 for K is recommended for estimating the generalization error. For an insightful discussion of the limitations of cross-validated choice among several evaluation methods, see Stone (1977).

5.5. Post-processing

Post-processing techniques can be utilized in most of the BCI designs to decrease the error rates. Some post-processing techniques can be designed specifically for a target application. For example, when a BCI system is used to activate a spelling device, some letters can be omitted without losing information. The system can also take into consideration the conditional probabilities of letters provided by one or two preceding letters and make corresponding suggestions to the patient (Kubler *et al* 1999). Such techniques may also be feasible for other applications and consequently increase the performance of the BCI systems.

There is a possibility that just after the end of a trial, some features of the brain signal reveal whether or not the trial was successful (that is, whether the outcome was or was not what the subject desired). These features are referred to as error potentials and can be used to detect errors in a BCI system and void the outcome. This error detection approach was encouraged by evidence that errors in conventional motor performances have detectable effects on the EEG recorded just after the error occurs (Falkenstein *et al* 1995, Falkenstein *et al* 2001, Gehring *et al* 1995). Whatever the nature of the error potential, the central decision for a BCI is how useful the error potential can be in detecting errors in single trials, and thereby improving accuracy. While its signal-to-noise ratio (SNR) is low, the error potential can improve the performance of a BCI system. In the meantime, better methods for recognizing and measuring the error potential could substantially improve its SNR, and thereby increase its impact on accuracy of a

BCI system. Such error potentials have been used in a few BCI systems to increase the performance (Bayliss *et al* 2004, Blankertz *et al* 2002b, 2003, Parra *et al* 2003b, Schalk *et al* 2000).

Another useful technique in decreasing false activations of BCI systems is to consider a measure of confidence in classification. In such a case, the output of the system can only be activated when the probability of the output being in an active state is greater than a given probability threshold or some criterion. Otherwise, the response of the BCI is considered ‘unknown’ and rejected to avoid making risky decisions. This is a useful way of reducing false decisions of the system (e.g., Cincotti *et al* (2003b), Millan *et al* (1998) and Penny *et al* (2000)) and might be used in any BCI design.

Considering mechanisms like debouncing the output of BCI designs also can reduce the number of false activations (Bashashati *et al* 2005, Borisoff *et al* 2004, Fatourehchi *et al* 2004, 2005, Muller-Putz *et al* 2005b, Obeid and Wolf 2004, Pfurtsceller *et al* 2005, Townsend *et al* 2004). These methods are specifically useful for so called asynchronous (self-paced) BCIs. Since false positives could happen in periods longer than just a few samples, using the debouncing technique in a manner similar to the debouncing of physical switches is expected to improve false activation rates (but with a cost in decreased re-activation time). The debounce component continuously monitors the output of the classifier. After an activation is detected (e.g. a change in logical state from ‘0’ to ‘1’ in a binary classifier), the output is activated for one time sample, then the output is forced to an inactive state for $T_d - 1$ time samples, where T_d is the debounce time period in samples. In some studies this time period is referred to as refractory period. As the debounce period is increased, the false activation rate is decreased for a given true positive rate. However, with increasing this time period, the re-activation time of the BCI system is impacted. The trade-off is clear and one needs to consider this for a given application.

6. Conclusions

We have completed the first comprehensive survey of signal processing methods used in BCI studies and published prior to January 2006. The results of this survey form a valuable and historical cross-reference for methods used in the following signal processing components of a BCI design: (1) pre-processing (signal enhancement), (2) feature selection/dimensionality reduction, (3) feature extraction, (4) feature classification and (5) post-processing methods. This survey shows which signal processing techniques have received more attention and which have not. This information is also valuable for newcomers to the field, as they can now find out which signal processing methods have been used for a certain type of a BCI system.

Many signal processing methods have been proposed and implemented in various brain-computer interfaces and comparison of these methods for different BCI applications

would be a useful task. However, at this point we cannot perform this task given the diversity of brain–computer interface systems from different aspects such as target application, neuromechanism used, amount of data tested, number of subjects and the amount of training they have received, recording systems, and experimental paradigms. Also acknowledged in McFarland *et al* (2006), we think that comparison of methods would be possible in well-designed systematic studies (Jackson *et al* 2006) and on established datasets such as the BCI Competition datasets (Blankertz *et al* 2004, 2006). We believe that, for a fair comparison of methods, more data would be needed as comparing methods with the data of one or two subjects does not necessarily guarantee the same findings on a larger subject pool. Jackson *et al* (2006) have provided a step toward this goal by proposing some ways to establish a systematic study both in design and reporting the results and we think that this task would only be possible with the collective help of all the researchers in this field.

We hope that this study will spawn further discussion of signal processing schemes for BCI designs. The proposed taxonomy and classes defined in tables 3–8 represent a proposed set of subcategories, not a final one, and we encourage others to revise or expand upon this initial set. Our direction in the future is to establish an online public-accessible database where research groups will be able to submit their signal processing designs as well as propose revisions/expansions of the proposed definitions and categories presented in this paper.

Appendix A. Index of terms

Index term	Description
AAR	Adaptive auto-regressive
AEP	Auditory evoked potential
AGR	Adaptive Gaussian representation
ALN	Adaptive logic network
ANC	Activity of neural cells
ANN	Artificial neural networks
AR	Auto-regressive
ARTMAP	Adaptive resonance theory MAP
ARX	Auto-regressive with exogenous input
BPF	Band-pass filter
C4.5	–
CAR	Common average referencing
CBR	Changes in brain rhythms
CCTM	Cross-correlation-based template matching
CER	Coarse-grained entropy rate
CHMM	Coupled hidden Markov model
CN2	–
CSP	Common spatial patterns
CSSD	Common spatial subspace decomposition
CSSP	Common spatio-spectral patterns
CTFR	Correlative time–frequency representation
CTFSR	Correlative time–frequency–space representation
DFT	Discrete Fourier transform
DSLQ	Distinctive sensitive learning vector quantization
ERD	Event-related desynchronization
ERN	Event-related negativity

Index term	Description
ERS	Event-related synchronization
FLD	Fisher’s linear discriminant
FFT	Fast Fourier transform
Freq-Norm	Frequency normalization
GA	Genetic algorithm
GAM	Generalized additive models
GLA	Generalized linear models
GPER	Gaussian process entropy rates
HMM	Hidden Markov model
ICA	Independent component analysis
IFFT	Inverse fast Fourier transform
k-NN	k-nearest neighbor
LDA	Linear discriminant analysis
LDS	Linear dynamical system
LGM	Linear Gaussian models implemented by Kalman filter
LMS	Least mean square
LPC	Linear predictive coding
LPF	Low-pass filter
LRP	Lateralized readiness potential
LVQ	Learning vector quantization
MD	Mahalanobis distance
MLP	Multi-layer perceptron neural networks
MN	Multiple neuromechanisms
MNF	Maximum noise fraction
MRA	Movement-related activity
NID3	–
NMF	Non-negative matrix factorization
NN	Neural networks
OLS1	Orthogonal least square
OPM	Outlier processing method
PCA	Principal component analysis (a.k.a. Karhounen Loeve transform)
PLV	Phase locking values
PPM	Piecewise Prony method
PSD	Power-spectral density
RBF	Radial basis function
RFE	Recursive feature/channel elimination
RNN	Recurrent neural network
SA-UK	Successive averaging and/or considering choice of unknown
SCP	Slow cortical potentials
SE	Spectral-entropy
SFFS	Sequential forward feature selection
SL	Surface Laplacian
SOFNN	Self-organizing feature neural network
SOM	Self-organizing map
SSEP	Somatosensory evoked potential
SSP	Signal space projection
SSVEP	Steady state visual evoked potential
STD	Standard deviation
SVD	Singular value decomposition
SVM	Support vector machine
SVR	Support vector machine regression
SWDA	Stepwise discriminant analysis
TBNN	Tree-based neural network
TFR	Time–frequency representation
VEFD	Variable epoch frequency decomposition
VEP	Visual evoked potential
WE	Wavelet entropy
WK	Wiener–Khinchine
ZDA	Z-scale-based discriminant analysis

References

- Allison B Z and Pineda J A 2003 ERPs evoked by different matrix sizes: implications for a brain–computer interface (BCI) system *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 110–3
- Allison B Z and Pineda J A 2005 Effects of SOA and flash pattern manipulations on ERPs, performance, and preference: implications for a BCI system *Int. J. Psychophysiol.* **59** 127–40
- Anderson C W, Devulapalli S V and Stolz E A 1995a Signal classification with different signal representations *Proc. IEEE Workshop on Neural Networks for Signal Processing* pp 475–83
- Anderson C W, Stolz E A and Shamsunder S 1995b Discriminating mental tasks using EEG represented by AR models *Proc. 17th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Montreal)* pp 875–6
- Anderson C W, Stolz E A and Shamsunder S 1998 Multivariate autoregressive models for classification of spontaneous electroencephalographic signals during mental tasks *IEEE Trans. Biomed. Eng.* **45** 277–86
- Babiloni C, Babiloni F, Carducci F, Cappa S F, Cincotti F, Del Percio C, Miniussi C, Moretti D V, Rossi S, Sosta K and Rossini P M 2004 Human cortical responses during one-bit short-term memory. A high-resolution EEG study on delayed choice reaction time tasks *Clin. Neurophysiol.* **115** 161–70
- Babiloni F, Bianchi L, Semeraro F, Millan J R, Mourino J, Cattini A, Salinari S, Marciiani M G and Cincotti F 2001a Mahalanobis distance-based classifiers are able to recognize EEG patterns by using few EEG electrodes *Proc. 23rd Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Istanbul)* pp 651–4
- Babiloni F, Cincotti F, Bianchi L, Pirri G, Millan J R, Mourino J, Salinari S and Marciiani M G 2001b Recognition of imagined hand movements with low resolution surface laplacian and linear classifiers *Med. Eng. Phys.* **23** 323–8
- Babiloni F, Cincotti F, Lazzarini L, Millan J, Mourino J, Varsta M, Heikkonen J, Bianchi L and Marciiani M G 2000 Linear classification of low-resolution EEG patterns produced by imagined hand movements *IEEE Trans. Rehabil. Eng.* **8** 186–8
- Balbale U H, Higgins J E, Bement S L and Levine S P 1999 Multi-channel analysis of human event-related cortical potentials for the development of a direct brain interface *Proc. 21st Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society & Annual Fall meeting of the Biomedical Engineering Society (Atlanta, GA)* p 447
- Barreto A B, Taberner A M and Vicente L M 1996a Classification of spatio-temporal EEG readiness potentials towards the development of a brain–computer interface *Proc. IEEE Southeast Conf. (Raleigh, NC)* pp 99–102
- Barreto A B, Taberner A M and Vicente L M 1996b Neural network classification of spatio-temporal EEG readiness potentials *Proc. IEEE 15th Southern Biomedical Engineering Conf. (Dayton, OH)* pp 73–6
- Bashashati A, Ward R K and Birch G E 2005 A new design of the asynchronous brain–computer interface using the knowledge of the path of features *Proc. 2nd IEEE-EMBS Conf. on Neural Engineering (Arlington, VA)* pp 101–4
- Bashashati A, Ward R K, Birch G E, Hashemi M R and Khalilzadeh M A 2003 Fractal dimension-based EEG biofeedback system *Proc. 25th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Cancun, Mexico)* pp 2220–3
- Bayliss J D 2003 Use of the evoked potential P3 component for control in a virtual apartment *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 113–6
- Bayliss J D and Ballard D H 1999 Single trial P300 recognition in a virtual environment *Proc. Int. ICSC Symp. on Soft Computing in Biomedicine (Genova, Italy)*
- Bayliss J D and Ballard D H 2000a Recognizing evoked potentials in a virtual environment *Advances in Neural Information Processing Systems* vol 12, ed S A Solla, T K Leen and K R Müller (Cambridge, MA: MIT Press) pp 3–9
- Bayliss J D and Ballard D H 2000b A virtual reality testbed for brain–computer interface research *IEEE Trans. Rehabil. Eng.* **8** 188–90
- Bayliss J D, Inverso S A and Tentler A 2004 Changing the P300 brain–computer interface *Cyberpsychol. Behav.* **7** 694–704
- Birbaumer N, Ghanayim N, Hinterberger T, Iversen I, Kotchoubey B, Kubler A, Perelmouter J, Taub E and Flor H 1999 A spelling device for the paralyzed *Nature* **398** 297–8
- Birbaumer N, Kubler A, Ghanayim N, Hinterberger T, Perelmouter J, Kaiser J, Iversen I, Kotchoubey B, Neumann N and Flor H 2000 The thought translation device (TTD) for completely paralyzed patients *IEEE Trans. Rehabil. Eng.* **8** 190–3
- Birch G E 1988 Single trial EEG signal analysis using outlier information *PhD Thesis* The University of British Columbia, Vancouver, Canada
- Birch G E, Bozorgzadeh Z and Mason S G 2002 Initial on-line evaluations of the LF-ASD brain–computer interface with able-bodied and spinal-cord subjects using imagined voluntary motor potentials *IEEE Trans. Neural Syst. Rehabil. Eng.* **10** 219–24
- Birch G E, Mason S G and Borisoff J F 2003 Current trends in brain–computer interface research at the Neil Squire Foundation *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 123–6
- Black M J, Bienenstock E, Donoghue J P, Wu W and Gao Y 2003 Connecting brains with machines: the neural control of 2D cursor movement *Proc. 2nd IEEE-EMBS Conf. on Neural Engineering (Arlington, VA)* pp 580–3
- Blanchard G and Blankertz B 2004 BCI competition 2003–data set IIa: Spatial patterns of self-controlled brain rhythm modulations *IEEE Trans. Biomed. Eng.* **51** 1062–6
- Blankertz B, Curio G and Müller K R 2002a Classifying single trial EEG: Towards brain–computer interfacing *Advances in Neural Information Processing Systems* vol 14, ed T G Dietterich, S Becker and Z Ghahramani (Cambridge, MA: MIT Press) pp 157–64
- Blankertz B, Dornhege G, Schäfer C, Krepek R, Kolmogorov J, Müller K R, Kunzmann V, Losch F and Curio G 2003 Boosting bit rates and error detection for the classification of fast-paced motor commands based on single-trial EEG analysis *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 127–31
- Blankertz B, Müller K R, Curio G, Vaughan T M, Schalk G, Wolpaw J R, Schlögl A, Neuper C, Pfurtscheller G, Hinterberger T, Schroder M and Birbaumer N 2004 The BCI competition 2003: progress and perspectives in detection and discrimination of EEG single trials *IEEE Trans. Biomed. Eng.* **51** 1044–51
- Blankertz B, Müller K R, Krusienski D J, Schalk G, Wolpaw J R, Schlögl A, Pfurtscheller G, Millan J R, Schroder M and Birbaumer N 2006 The BCI competition: III. Validating alternative approaches to actual BCI problems *IEEE Trans. Neural Syst. Rehabil. Eng.* **14** 153–9
- Blankertz B, Schäfer C, Dornhege G and Curio G 2002b Single trial detection of EEG error potentials: a tool for increasing BCI transmission rates *Proc. Int. Conf. on Artificial Neural Networks (Madrid, Spain)* pp 1137–43
- Boostani R and Moradi M H 2004 A new approach in the BCI research based on fractal dimension as feature and adaboost as classifier *J. Neural Eng.* **1** 212–7
- Borisoff J F, Mason S G, Bashashati A and Birch G E 2004 Brain–computer interface design for asynchronous control applications: Improvements to the LF-ASD asynchronous brain switch *IEEE Trans. Biomed. Eng.* **51** 985–92
- Bostanov V 2004 BCI competition 2003–data sets Ib and IIb: feature extraction from event-related brain potentials with the

- continuous wavelet transform and the t-value scalogram *IEEE Trans. Biomed. Eng.* **51** 1057–61
- Bozorgzadeh Z, Birch G E and Mason S G 2000 The LF-ASD BCI: on-line identification of imagined finger movements in spontaneous EEG with able-bodied subjects *Proc. IEEE Int. Conf. on Acoustics, Speech and Signal Processing (Istanbul, Turkey)* pp 109–13
- Burke D, Kelly S, de Chazal P and Reilly R 2002 A simultaneous filtering and feature extraction strategy for direct brain interfacing *Proc. 24th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Houston, TX)* pp 279–80
- Burke D P, Kelly S P, de Chazal P, Reilly R B and Finucane C 2005 A parametric feature extraction and classification strategy for brain–computer interfacing *IEEE Trans. Neural Syst. Rehabil. Eng.* **13** 12–7
- Byron O, He J P, Hu J and Si J 2005 A conceptual brain machine interface system *Proc. 1st Int. Conf. on Neural Interface and Control (Wuhan, China)* pp 40–44
- Calhoun G L and McMillan G R 1996 EEG-based control for human–computer interaction *Proc. 3rd Symp. On Human Interaction with Complex Systems (Dayton, OH)* pp 4–9
- Carmena J M, Lebedev M A, Crist R E, O’Doherty J E, Santucci D M, Dimitrov D F, Patil P G, Henriquez C S and Nicolelis M A 2003 Learning to control a brain–machine interface for reaching and grasping by primates *Plos Biol.* **1** 193–208
- Carmena J M, Lebedev M A, Henriquez C S and Nicolelis M A 2005 Stable ensemble performance with single-neuron variability during reaching movements in primates *J. Neurosci.* **25** 10712–6
- Chapin J K, Moxon K A, Markowitz R S and Nicolelis M A 1999 Real-time control of a robot arm using simultaneously recorded neurons in the motor cortex *Nat. Neurosci.* **2** 664–70
- Cheng M and Gao S 1999 An EEG-based cursor control system *Proc. 21st Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society & Annual Fall meeting of the Biomedical Engineering Society (Atlanta, GA)* p 669
- Cheng M, Gao X, Gao S and Wang B 2005 Stimulation frequency extraction in SSVEP-based brain–computer interface *Proc. 1st Int. Conf. on Neural Interface and Control (Wuhan, China)* pp 64–7
- Cheng M, Gao X, Gao S and Xu D 2002 Design and implementation of a brain–computer interface with high transfer rates *IEEE Trans. Biomed. Eng.* **49** 1181–6
- Cheng M, Jia W, Gao X, Gao S and Yang F 2004 Mu rhythm-based cursor control: an offline analysis *Clin. Neurophysiol.* **115** 745–51
- Cheng M, Xu D, Gao X and Gao S 2001 Brain–computer interface with high transfer rates *Proc. 8th Int. Conf. on Neural Information Processing (Shanghai, China)*
- Cho B H, Kim S, Shin D I, Lee J H, Lee S M, Kim I Y and Kim S I 2004 Neurofeedback training with virtual reality for inattention and impulsiveness *Cyberpsychol. Behav.* **7** 519–26
- Cincotti F, Bianchi L, Millan J, Mourino J, Salinari S, Marciani M G and Babiloni F 2001 Brain–computer interface: the use of low resolution surface laplacian and linear classifiers for the recognition of imagined hand movements *Proc. 23rd Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Istanbul)* pp 655–8
- Cincotti F, Mattia D, Babiloni C, Carducci F, Salinari S, Bianchi L, Marciani M G and Babiloni F 2003a The use of EEG modifications due to motor imagery for brain–computer interfaces *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 131–3
- Cincotti F, Scipione A, Timperi A, Mattia D, Marciani M G, Millan J, Salinari S, Bianchi L and Babiloni F 2003b Comparison of different feature classifiers for brain–computer interfaces *Proc. 25th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Cancun, Mexico)* pp 645–7
- Costa E J and Cabral E F 2000 EEG-based discrimination between imagination of left and right hand movements using adaptive gaussian representation *Med. Eng. Phys.* **22** 345–8
- Coyle D, Prasad G and McGinnity T M 2004 Extracting features for a brain–computer interface by self-organising fuzzy neural network-based time series prediction *Proc. 26th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (San Francisco, CA)* pp 4371–4
- Coyle D, Prasad G and McGinnity T M 2005 A time-frequency approach to feature extraction for a brain–computer interface with a comparative analysis of performance measures *EURASIP J. Appl. Signal Process.* **2005** 3141–51
- Curran E, Sykacek P, Stokes M, Roberts S J, Penny W, Johnsrude I and Owen A M 2004 Cognitive tasks for driving a brain–computer interfacing system: a pilot study *IEEE Trans. Neural Syst. Rehabil. Eng.* **12** 48–54
- Darmanjian S, Kim S P, Nechbya M C, Morrison S, Principe J, Wessberg J and Nicolelis M A 2003 Bimodal brain–machine interface for motor control of robotic prosthetic *Proc. IEEE/RSJ Int. Conf. on Intelligent Robots and Systems (Las Vegas, NV)* pp 3612–7
- De Jong K A 1975 An analysis of the behaviour of a class of genetic adaptive systems *PhD Thesis* University of Michigan, Ann Arbor, MI (Diss. Abstr. Int. 36(10), 5140B, University Microfilms No. 76–9381. 44)
- Deecke L and Kornhuber H 1976 Voluntary finger movement in man: Cerebral potentials and theory *Biol. Cybern.* **23** 99–119
- Donchin E, Spencer K M and Wijesinghe R 2000 The mental prosthesis: assessing the speed of a P300-based brain–computer interface *IEEE Trans. Rehabil. Eng.* **8** 174–9
- Donoghue J P 2002 Connecting cortex to machines: recent advances in brain interfaces *Nat. Neurosci.* **5** 1085–8
- Dornhege G, Blankertz B and Curio G 2003 Speeding up classification of multi-channel brain–computer interfaces: common spatial patterns for slow cortical potentials *Proc. 2nd IEEE-EMBS Conf. on Neural Engineering (Arlington, VA)* pp 595–8
- Dornhege G, Blankertz B, Curio G and Muller K R 2004 Boosting bit rates in noninvasive EEG single-trial classifications by feature combination and multiclass paradigms *IEEE Trans. Biomed. Eng.* **51** 993–1002
- Erfanian A and Erfani A 2004 ICA-based classification scheme for EEG-based brain–computer interface: the role of mental practice and concentration skills *Proc. 26th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (San Francisco, CA)* pp 235–8
- Fabiani G E, McFarland D J, Wolpaw J R and Pfurtscheller G 2004 Conversion of EEG activity into cursor movement by a brain–computer interface (BCI) *IEEE Trans. Neural Syst. Rehabil. Eng.* **12** 331–8
- Falkenstein M, Hielscher H, Dziobek I, Schwarzenau P, Hoormann J, Sunderman B and Hohnsbein J 2001 Action monitoring, error detection, and the basal ganglia: an ERP study *Neuroreport* **12** 157–61
- Falkenstein M, Hohnsbein J and Hoormann J 1995 Event-related potential correlates of errors in reaction tasks *Perspectives of Event-Related Brain Potentials Research* ed Karmos *et al* (Amsterdam: Elsevier) pp 287–96
- Farwell L A and Donchin E 1988 Talking off the top of your head: Towards a mental prosthesis utilizing event-related brain potentials *Electroencephalogr. Clin. Neurophysiol.* **80** 510–23
- Fatourechi M, Bashashati A, Borisoff J F, Birch G E and Ward R K 2004 Improving the performance of the LF-ASD brain–computer interface by means of genetic algorithm *Proc. IEEE Symp. on Signal Processing and Information Technology (Rome, Italy)* pp 38–41
- Fatourechi M, Bashashati A, Ward R K and Birch G E 2005 A hybrid genetic algorithm approach for improving the performance of the LF-ASD brain–computer interface *Proc.*

- IEEE Int. Conf. on Acoustics, Speech and Signal Processing (Philadelphia, PA)* pp 345–8
- Flotzinger D, Pregezer M and Pfurtscheller G 1994 Feature selection with distinction sensitive learning vector quantisation and genetic algorithms *Proc. IEEE Int. Conf. on Neural Networks (Orlando, FL)* pp 3448–51
- Fukada S, Tatsumi D, Tsujimoto H and Inokuchi S 1998 Studies of input speed of word inputting system using event-related potential *Proc. 20th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Hong Kong)* pp 1458–60
- Gao Y, Black M J, Bienenstock E, Shoham S and Donoghue J P 2002 Probabilistic inference of hand motion from neural activity in motor cortex *Advances in Neural Information Processing Systems* vol 14, ed T G Dietterich, S Becker and Z Ghahramani (Cambridge, MA: MIT Press) pp 213–20
- Gao Y, Black M J, Bienenstock E, Wu W and Donoghue J P 2003a A quantitative comparison of linear and non-linear models of motor cortical activity for the encoding and decoding of arm motions *Proc. 25th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Hong Kong)* pp 189–92
- Gao X, Xu D, Cheng M and Gao S 2003b A BCI-based environmental controller for the motion-disabled *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 137–40
- Gao X, Xu N, Hong B, Gao S and Yang F 2004 Optimal selection of independent components for event-related brain electrical potential enhancement *Proc. IEEE Int. Workshop on Biomedical Circuits and Systems (Singapore)* pp S3/5/INV-S3/5/1-4
- Garcia G and Ebrahimi T 2002 Time–frequency–space kernel for single EEG-trial classification *Presented at the 5th Nordic Signal Processing Symp. (Norway)*
- Garcia G, Ebrahimi T and Vesin J M 2002 Classification of EEG signals in the ambiguity domain for brain–computer interface applications *Proc. IEEE 14th Int. Conf. on Digital Signal Processing (Santorini, Greece)* pp 301–305
- Garcia G, Ebrahimi T and Vesin J M 2003a Correlative exploration of EEG signals for direct brain–computer communications *Proc. IEEE Int. Conf. on Acoustics, Speech, and Signal Processing (Hong Kong)* pp 816–9
- Garcia G, Ebrahimi T and Vesin J M 2003b Support vector EEG classification in the fourier and time-frequency correlation domains *Proc. 1st IEEE-EMBS Conf. on Neural Engineering (Capri Island, Italy)* pp 591–4
- Garcia G N, Ebrahimi T, Vesin J M and Villca A 2003c Direct brain–computer communication with user rewarding mechanism *Proc. IEEE Int. Symp. in Information Theory (Yokohama, Japan)* p 221
- Garrett D, Peterson D A, Anderson C W and Thaut M H 2003 Comparison of linear, nonlinear, and feature selection methods for EEG signal *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 141–4
- Gehring W J, Coles M G H, Meyer D E and Donchin E 1995 A brain potential manifestation of error-related processing *Perspectives of Event-Related Brain Potentials Research (EEG Suppl vol 44)* ed Karmos *et al* (Amsterdam:Elsevier) pp 261–72
- Georgopoulos A P, Langheim F J, Leuthold A C and Merkle A N 2005 Magnetoencephalographic signals predict movement trajectory in space *Exp. Brain Res.* **167** 132–5
- Gevins A, Bressler S L, Morgan N H, Cutillo B A, White R M, Greer D S and Illes J 1989 Event-related covariances during a bimanual visuomotor task: I. Methods and analysis of stimulus- and response-locked data *Electroencephalogr. Clin. Neurophysiol.* **74** 58–75
- Glassman E L 2005 A wavelet-like filter based on neuron action potentials for analysis of human scalp electroencephalographs *IEEE Trans. Biomed. Eng.* **52** 1851–62
- Graimann B, Huggins J E, Levine S P and Pfurtscheller G 2003a Detection of ERP and ERD/ERS patterns in single ECoG channels *Proc. 1st IEEE-EMBS Conf. on Neural Engineering (Capri Island, Italy)* pp 614–7
- Graimann B, Huggins J E, Levine S P and Pfurtscheller G 2004 Toward a direct brain interface based on human subdural recordings and wavelet-packet analysis *IEEE Trans. Biomed. Eng.* **51** 954–62
- Graimann B, Huggins J E, Schlogl A, Levine S P and Pfurtscheller G 2003b Detection of movement-related desynchronization patterns in ongoing single-channel electrocorticogram *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 276–81
- Guan J, Chen Y, Lin J, Yun Y and Huang M 2005 N2 components as features for brain–computer interface *Proc. 1st Int. Conf. on Neural Interface and Control (Wuhan, China)* pp 45–9
- Guan C, Thulasidas M and Wu J 2004 High performance P300 speller for brain–computer interface *Proc. IEEE Int. Workshop on Biomedical Circuits and Systems (Singapore)* pp S3/5/INV-S3/13–16
- Guger C, Edlinger G, Harkam W, Niedermayer I and Pfurtscheller G 2003a How many people are able to operate an EEG-based brain–computer interface (BCI)? *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 145–7
- Guger C, Edlinger G and Pfurtscheller G 2003b How many people are able to operate an EEG-based brain–computer interface (BCI)? Presented at the *2nd Int. Meeting on Brain–Computer Interfaces for Communication and Control (Albany, NY)*
- Guger C, Muller G, Neuper C, Krausz G, Niedermayer I and Pfurtscheller G 2000a Brain–computer communication system: the EEG-based control of a hand orthosis in a quadriplegic patient Presented at the *Int. Conf. on Computers Helping People with Special Needs (Karlsruhe, Germany)*
- Guger C, Ramoser H and Pfurtscheller G 2000b Real-time EEG analysis with subject-specific spatial patterns for a brain–computer interface (BCI) *IEEE Trans. Rehabil. Eng.* **8** 447–56
- Guger C, Schloegl A, Walterspercher D and Pfurtscheller G 1999 Design of an EEG-based brain–computer interface (BCI) from standard components running in real-time under windows *Biomed. Tech. (Berl)* **44** 12–6
- Gysels E and Celka P 2004 Phase synchronization for the recognition of mental tasks in a brain–computer interface *IEEE Trans. Neural Syst. Rehabil. Eng.* **12** 406–15
- Gysels E, Renevey P and Celka P 2005 SVM-based recursive feature elimination to compare phase synchronization computed from broadband and narrowband EEG signals in brain–computer interfaces *Signal Process.* **85** 2178–89
- Hallett M 1994 Movement-related cortical potentials *Electromyogr. Clin. Neurophysiol.* **34** 5–13
- Haselsteiner E and Pfurtscheller G 2000 Using time-dependent neural networks for EEG classification *IEEE Trans. Rehabil. Eng.* **8** 457–63
- Hatsopoulos N, Joshi J and O’Leary J G 2004 Decoding continuous and discrete motor behaviors using motor and premotor cortical ensembles *J. Neurophysiol.* **92** 1165–74
- Hill N J, Lal T N, Bierig K, Birbaumer N and Scholkopf B 2004 Attention modulation of auditory event-related potentials in a brain–computer interface *Proc. IEEE Int. Workshop on Biomedical Circuits and Systems (Singapore)* pp S3/5/INV-S3/17–20
- Hinterberger T and Baier G 2005 Parametric orchestral sonification of EEG in real time *IEEE Multimedia* **12** 70–9
- Hinterberger T, Kubler A, Kaiser J, Neumann N and Birbaumer N 2003 A brain–computer interface (BCI) for the locked-in: comparison of different EEG classifications for the thought translation device *Electroencephalogr. Clin. Neurophysiol.* **114** 416–25
- Hinterberger T, Neumann N, Pham M, Kubler A, Grether A, Hofmayer N, Wilhelm B, Flor H and Birbaumer N 2004a

- A multimodal brain-based feedback and communication system *Exp. Brain Res.* **154** 521–6
- Hinterberger T, Veit R, Wilhelm B, Weiskopf N, Vatine J J and Birbaumer N 2005a Neuronal mechanisms underlying control of a brain–computer interface *Eur. J. Neurosci.* **21** 3169–81
- Hinterberger T, Weiskopf N, Veit R, Wilhelm B, Betta E and Birbaumer N 2004b An EEG-driven brain–computer interface combined with functional magnetic resonance imaging (fMRI) *IEEE Trans. Biomed. Eng.* **51** 971–4
- Hinterberger T, Wilhelm B, Mellinger J, Kotchoubey B and Birbaumer N 2005b A device for the detection of cognitive brain functions in completely paralyzed or unresponsive patients *IEEE Trans. Biomed. Eng.* **52** 211–20
- Hu J, Si J, Olson B P and He J 2004 Principle component feature detector for motor cortical control *Proc. 26th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (San Francisco, CA)* pp 4021–4
- Huan N J and Palaniappan R 2004 Neural network classification of autoregressive features from electroencephalogram signals for brain–computer interface design *J. Neural Eng.* **1** 142–50
- Huan N J and Palaniappan R 2005 Classification of mental tasks using fixed and adaptive autoregressive models of EEG signals *Proc. 27th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Shanghai, China)* pp 633–6
- Huggins J E *et al* 2003 Electrooculogram as the basis for a direct brain interface: Opportunities for improved detection accuracy *Proc. 1st IEEE-EMBS Conf. on Neural Engineering (Capri Island, Italy)* pp 587–90
- Huggins J E, Levine S P, Bement S L, Kushwaha R K, Schuh L A, Passaro E A, Rohde M M, Ross D A, Elisevich K V and Smith B J 1999 Detection of event-related potentials for development of a direct brain interface *J. Clin. Neurophysiol.* **16** 448–55
- Hung C I, Lee P L, Wu Y T, Chen L F, Yeh T C and Hsieh J C 2005 Recognition of motor imagery electroencephalography using independent component analysis and machine classifiers *Ann. Biomed. Eng.* **33** 1053–70
- Isaacs R E, Weber D J and Schwartz A B 2000 Work toward real-time control of a cortical neural prosthesis *IEEE Trans. Rehabil. Eng.* **8** 196–8
- Ivanova I, Pfurtscheller G and Andrew C 1995 AI-based classification of single-trial EEG data *Proc. 17th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Montreal, Canada)* pp 703–4
- Jackson M M, Mason S G and Birch G E 2006 Analyzing trends in brain interface technology: a method to compare studies *Ann. Biomed. Eng.* **34** 859–78
- Jansen B H, Allam A, Kota P, Lachance K, Osho A and Sundaresan K 2004 An exploratory study of factors affecting single trial P300 detection *IEEE Trans. Biomed. Eng.* **51** 975–8
- Jasper H and Penfield W 1949 Electrooculograms in man: effect of voluntary movement upon the electrical activity of the precentral gyrus *Arch. Psychiat. Nervenkr.* **183** 163–74
- Jia W, Zhao X, Liu H, Gao X, Gao S and Yang F 2004 Classification of single trial EEG during motor imagery based on ERD *Proc. 26th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (San Francisco, CA)* pp 5–8
- Kaiser J, Kubler A, Hinterberger T, Neumann N and Birbaumer N 2002 A non-invasive communication device for the paralyzed *Minim. Invasive Neurosurg.* **45** 19–23
- Kaiser J, Perelmouter J, Iversen I H, Neumann N, Ghanayim N, Hinterberger T, Kubler A, Kotchoubey B and Birbaumer N 2001 Self-initiation of EEG-based communication in paralyzed patients *Electroencephalogr. Clin. Neurophysiol.* **112** 551–4
- Kalcher J, Flotzinger D and Pfurtscheller G 1992 A new approach to a brain–computer-interface (BCI) based on learning vector quantization (LVQ3) *Proc. 14th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Paris, France)* pp 1658–9
- Kalcher J, Flotzinger D and Pfurtscheller G 1993 Graz brain–computer interface: an EEG-based cursor control system *Proc. 15th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (San Diego, CA)* pp 1264–5
- Kaper M, Meinicke P, Grossekhoefer U, Lingner T and Ritter H 2004 BCI competition 2003–data set IIb: support vector machines for the P300 speller paradigm *IEEE Trans. Biomed. Eng.* **51** 1073–6
- Kaper M and Ritter H 2004a Generalizing to new subjects in brain–computer interfacing *Proc. 26th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (San Francisco, CA)* pp 4363–6
- Kaper M and Ritter H 2004b Progress in P300-based brain–computer interfacing *Proc. IEEE Int. Workshop on Biomedical Circuits and Systems (Singapore)* pp S3/5/INV-S3/59–12
- Karniel A, Fleming K M, Sanguineti V, Alford S and Mussa-Ivaldi F A 2002 Dynamic properties of the lamprey’s neuronal circuits as it drives a two-wheeled robot *Proc. SAB’2002 Workshop on Motor Control in Humans and Robots: on the Interplay of Real Brains and Artificial Devices (Edinburgh, Scotland)* pp 29–36
- Kawakami T, Inoue M, Kobayashi Y and Nakashima K 1996 Application of event related potentials to communication aids *Proc. 18th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Amsterdam, Holland)* pp 2229–31
- Keirn Z A and Aunon J I 1990 A new mode of communication between man and his surroundings *IEEE Trans. Biomed. Eng.* **37** 1209–14
- Kelly S, Burke D, Chazal P d and Reilly R 2002a Parametric models and classification for direct brain interfaces *Proc. Irish Signal and Systems Conf. (Cork, Ireland)* pp 25–6
- Kelly S, Burke D, Chazal P d and Reilly R 2002b Parametric models and spectral analysis for classification in brain–computer interfaces *Proc. 14th IEEE Int. Conf. on Digital Signal Processing (Santorini, Greece)* pp 307–10
- Kelly S P, Lalor E, Finucane C and Reilly R B 2004 A comparison of covert and overt attention as a control option in a steady-state visual evoked potential-based brain–computer interface *Proc. 26th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (San Francisco, CA)* pp 4725–8
- Kelly S P, Lalor E, Reilly R B and Foxe J J 2005a Independent brain–computer interface control using visual spatial attention-dependent modulations of parieto-occipital alpha *Proc. IEEE-EMBS Conf. on Neural Engineering (Arlington, VA)* pp 667–70
- Kelly S P, Lalor E C, Finucane C, McDarby G and Reilly R B 2005b Visual spatial attention control in an independent brain–computer interface *IEEE Trans. Biomed. Eng.* **52** 1588–96
- Kelly S P, Lalor E C, Reilly R B and Foxe J J 2005c Visual spatial attention tracking using high-density SSVEP data for independent brain–computer communication *IEEE Trans. Neural Syst. Rehabil. Eng.* **13** 172–8
- Kemere C, Shenoy K V and Meng T H 2004 Model-based neural decoding of reaching movements: a maximum likelihood approach *IEEE Trans. Biomed. Eng.* **51** 925–32
- Kennedy P R, Bakay R A, Moore M M, Adams K and Goldwithe J 2000 Direct control of a computer from the human central nervous system *IEEE Trans. Rehabil. Eng.* **8** 198–202
- Kim K H, Kim S S and Kim S J 2005a Superiority of nonlinear mapping in decoding multiple single-unit neuronal spike trains: a simulation study *J. Neurosci. Methods* **150** 202–11
- Kim S P, Rao Y N, Erdogmus D, Sanchez J C, Nicoletis M A L and Principe J C 2005b Determining patterns in neural activity for reaching movements using nonnegative matrix factorization *EURASIP J. Appl. Signal Process.* **19** 3113–21

- Kira K and Rendell L A 1992 The feature selection problem: traditional methods and a new algorithm *Proc. of the 10th National Conf. on Artificial Intelligence (San Jose, CA)* pp 129–34
- Kositsky M, Karniel A, Alford S, Fleming K M and Mussa-Ivaldi F A 2003 Dynamical dimension of a hybrid neurobotic system *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 155–9
- Kostov A and Polak M 1997 Prospects of computer access using voluntary modulated EEG signal *Proc. ECPD Symposium on Brain & Consciousness (Belgrade, Yugoslavia)* pp 233–6
- Kostov A and Polak M 2000 Parallel man-machine training in development of EEG-based cursor control *IEEE Trans. Rehabil. Eng.* **8** 203–5
- Kozelka J W and Pedley T A 1990 Beta and mu rhythms *J. Clin. Neurophysiol.* **7** 191–207
- Krauledat M, Dornhege G, Blankertz B, Losch F, Curio G and Muller K 2004 Improving speed and accuracy of brain–computer interfaces using readiness potential features *Proc. 26th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (San Francisco, CA)* pp 4511–5
- Krausz G, Scherer R, Korisek G and Pfurtscheller G 2003 Critical decision-speed and information transfer in the ‘graz brain–computer interface’ *Appl. Psychophysiol. Biofeedback* **28** 233–40
- Kubler A, Kotchoubey B, Hinterberger T, Ghanayim N, Perelmouter J, Schauer M, Fritsch C, Taub E and Birbaumer N 1999 The thought translation device: a neurophysiological approach to communication in total motor paralysis *Exp. Brain Res.* **124** 223–2
- Kubler A, Kotchoubey B, Kaiser J, Wolpaw J R and Birbaumer N 2001a Brain–computer communication: unlocking the locked *Psychol. Bull.* **127** 358–75
- Kubler A, Kotchoubey B, Salzmann H P, Ghanayim N, Perelmouter J, Homberg V and Birbaumer N 1998 Self-regulation of slow cortical potentials in completely paralyzed human patients *Neurosci. Lett.* **252** 171–4
- Kubler A, Neumann N, Kaiser J, Kotchoubey B, Hinterberger T and Birbaumer N P 2001b Brain–computer communication: Self-regulation of slow cortical potentials for verbal communication *Arch. Phys. Med. Rehabil.* **82** 1533–9
- Kubler A, Nijboer F, Mellinger J, Vaughan T M, Pawelzik H, Schalk G, McFarland D J, Birbaumer N and Wolpaw J R 2005 Patients with ALS can use sensorimotor rhythms to operate a brain–computer interface *Neurology* **64** 1775–7
- LaCourse J R and Wilson E 2003 Brainiac: a brain computer link *Proc. IEEE Instrumentation and Measurement Technology Conf (Vail, CO)* pp 587–92
- Lal T N, Schroder M, Hinterberger T, Weston J, Bogdan M, Birbaumer N and Scholkopf B 2004 Support vector channel selection in BCI *IEEE Trans. Biomed. Eng.* **51** 1003–10
- Lalor E, Kelly S, Finucane C, Burke R, Smith R, Reilly R and McDarby G 2005 Steady-state VEP-based brain–computer interface control in an immersive 3-D gaming environment *EURASIP J. Appl. Signal Process.* **19** 3156–64
- Laubach M, Wessberg J and Nicolelis M A 2000 Cortical ensemble activity increasingly predicts behaviour outcomes during learning of a motor task *Nature* **405** 567–71
- Lebedev M A, Carmena J M, O’Doherty J E, Zacksenhouse M, Henriquez C S, Principe J C and Nicolelis M A 2005 Cortical ensemble adaptation to represent velocity of an artificial actuator controlled by a brain–machine interface *J. Neurosci.* **25** 4681–93
- Lee H and Choi S 2002 PCA-based linear dynamical systems for multichannel EEG classification *Proc. 9th Int. Conf. on Neural Information Processing (Singapore)* pp 745–9
- Lee H and Choi S 2003 PCA+HMM+SVM for EEG pattern classification *Proc. IEEE 7th Int. Symp. on Signal Processing and Its Applications (Paris, France)* pp 541–4
- Lee P L, Wu C H, Hsieh J C and Wu Y T 2005 Visual evoked potential actuated brain–computer interface: a brain-actuated cursor system *Electron. Lett.* **41** 832–4
- Leeb R and Pfurtscheller G 2004 Walking through a virtual city by thought *Proc. 26th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (San Francisco, CA)* pp 4503–6
- Lemm S, Blankertz B, Curio G and Muller K R 2005 Spatio-spectral filters for improving the classification of single trial EEG *IEEE Trans. Biomed. Eng.* **52** 1541–8
- Lemm S, Schafer C and Curio G 2004 BCI competition 2003–data set III: probabilistic modeling of sensorimotor mu rhythms for classification of imaginary hand movements *IEEE Trans. Biomed. Eng.* **51** 1077–80
- Levine S P, Huggins J E, Bement S L, Kushwaha R K, Schuh L A, Passaro E A, Rohde M M and Ross D A 1999 Identification of electrocorticogram patterns as the basis for a direct brain interface *J. Clin. Neurophysiol.* **16** 439–47
- Levine S P, Huggins J E, Bement S L, Kushwaha R K, Schuh L A, Rohde M M, Passaro E A, Ross D A, Elisevich K V and Smith B J 2000 A direct brain interface based on event-related potentials *IEEE Trans. Rehabil. Eng.* **8** 180–5
- Li Y, Cichocki A, Guan C and Qin J 2004a Sparse factorization preprocessing-based offline analysis for a cursor control experiment *Proc. IEEE Int. Workshop on Biomedical Circuits and Systems (Singapore)* pp S3/5/INV-S3/5/5–8
- Li Y, Gao X, Liu H and Gao S 2004b Classification of single-trial electroencephalogram during finger movement *IEEE Trans. Biomed. Eng.* **51** 1019–25
- Lisogurski D and Birch G E 1998 Identification of finger flexions from continuous EEG as a brain–computer interface *Proc. 20th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Hong Kong)* pp 2004–7
- Liu H S, Gao X, Yang F and Gao S 2003 Imagined hand movement identification based on spatio-temporal pattern recognition of EEG *Proc. IEEE-EMBS Conf. on Neural Engineering (Capri Island, Italy)* pp 599–602
- Liu H, Wang J and Zheng C 2005 Mental tasks classification and their EEG structures analysis by using the growing hierarchical self-organizing map *Proc. 1st Int. Conf. on Neural Interface and Control (Wuhan, China)* pp 115–8
- Mahmoudi B and Erfanian A 2002 Single-channel EEG-based prosthetic hand grasp control for amputee subjects *Proc. 24th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Houston, TX)* pp 2406–7
- Mason S G and Birch G E 2000 A brain-controlled switch for asynchronous control applications *IEEE Trans. Biomed. Eng.* **47** 1297–307
- Mason S G and Birch G E 2003 A general framework for brain–computer interface design *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 70–85
- Mason S G, Bohringer R, Borisoff J F and Birch G E 2004 Real-time control of a video game with a direct brain–computer interface *J. Clin. Neurophysiol.* **21** 404–8
- McFarland D J, Anderson C W, Muller K R, Schlögl A and Krusienski D J 2006 BCI meeting 2005—workshop on BCI signal processing: feature extraction and translation *IEEE Trans. Neural Syst. Rehabil. Eng.* **14** 135–8
- McFarland D J, McCane L M, David S V and Wolpaw J R 1997 Spatial filter selection for EEG-based communication *Electroencephalogr. Clin. Neurophysiol.* **103** 386–94
- McFarland D J, Sarnacki W A, Vaughan T M and Wolpaw J R 2005 Brain–computer interface (BCI) operation: signal and noise during early training sessions *Clin. Neurophysiol.* **116** 56–62
- McFarland D J, Sarnacki W A and Wolpaw J R 2003 Brain–computer interface (BCI) operation: optimizing information transfer rates *Biol. Psychol.* **63** 237–51

- McFarland D and Wolpaw J R 1998 EEG-based communication and control: short-term role of feedback *IEEE Trans. Rehabil. Eng.* **6** 7–11
- McMillan G R and Calhoun G L 1995 Direct brain interface utilizing self-regulation of steady-state visual evoked response (SSVER) *Proc. RESNA '95 Annual Conf. (Vancouver, BC)* pp 693–5
- Mensh B D, Werfel J and Seung H S 2004 BCI competition 2003—data set ia: combining gamma-band power with slow cortical potentials to improve single-trial classification of electroencephalographic signals *IEEE Trans. Biomed. Eng.* **51** 1052–6
- Middendorf M, McMillan G, Calhoun G and Jones K S 2000 Brain–computer interfaces based on the steady-state visual-evoked response *IEEE Trans. Rehabil. Eng.* **8** 211–4
- Millan J R 2004 On the need for on-line learning in brain–computer interfaces *Proc. Annual Int. Joint Conf. on Neural Networks (Budapest, Hungary)*
- Millan J R, Renkens F, Mourino J and Gerstner W 2004a Noninvasive brain-actuated control of a mobile robot by human EEG *IEEE Trans. Biomed. Eng.* **51** 1026–33
- Millan J R, Renkens F, Mourino J and Gerstner W 2003a Non-invasive brain-actuated control of a mobile robot *Proc. 18th Int. Joint Conf. on Artificial Intelligence (Acapulco, Mexico)* pp 1121–6
- Millan J R, Renkens F, Mourino J and Gerstner W 2004b Brain-actuated interaction *Artif. Intell.* **159** 241–59
- Millan J, Franze M, Mourino J, Cincotti F and Babiloni F 2002a Relevant EEG features for the classification of spontaneous motor-related tasks *Biol. Cybern.* **86** 89–95
- Millan J, Mourino J, Franze M, Cincotti F, Varsta M, Heikkinen J and Babiloni F 2002b A local neural classifier for the recognition of EEG patterns associated to mental tasks *IEEE Trans. Neural Netw.* **13** 678–86
- Millan J R, Mourino J, Babiloni F, Cincotti F, Varsta M and Heikkinen J 2000a Local neural classifier for EEG-based recognition of mental tasks *Proc. IEEE-INNS-ENNS Int. Joint Conf. on Neural Networks (Como, Italy)* pp 632–6
- Millan J R, Mourino J, Cincotti F, Varsta M, Heikkinen J, Topani F, Marciani M G, Kaski K and Babiloni F 2000b Neural networks for robust classification of mental tasks *Proc. 22nd Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Chicago, IL)* pp 1380–2
- Millan J R, Mourino J, Marciani M G, Babiloni F, Topani F, Canale I, Heikkinen J and Kaski K 1998 Adaptive brain interfaces for physically-disabled people *Proc. 20th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Hong Kong)* pp 2008–11
- Millan J R and Mourino J 2003b Asynchronous BCI and local neural classifiers: an overview of the adaptive brain interface project *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 159–61
- Molina G G, Ebrahimi T and Vesin J M 2003 Joint time-frequency-space classification of EEG in a brain–computer interface application *EURASIP J. Appl. Signal Process.* **7** 713–29
- Mourino J, Chiappa S, Jane R and Millan J R 2002 Evolution of the mental states operating a brain–computer interface *Proc. of the Int. Federation for Medical and Biological Engineering (Vienna, Austria)* pp 600–1
- Muller K R, Anderson C W and Birch G E 2003a Linear and nonlinear methods for brain–computer interfaces *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 165–9
- Muller K R, Curio G, Blankertz B and Dornhege G 2003b Combining features for BCI *Advances in Neural Information Processing Systems (NIPS02)* vol 15, ed S Becker, S Thrun and K Obermayer pp 1115–22
- Muller G R, Neuper C and Pfurtscheller G 2003c Implementation of a telemonitoring system for the control of an EEG-based brain–computer interface *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 54–9
- Muller-Putz G R, Scherer R, Braunedis C and Pfurtscheller G 2005a Steady-state visual evoked potential (SSVEP)-based communication: impact of harmonic frequency components *J. Neural Eng.* **2** 123–30
- Muller-Putz G R, Scherer R, Pfurtscheller G and Rupp R 2005b EEG-based neuroprosthesis control: a step towards clinical practice *Neurosci. Lett.* **382** 169–74
- Musallam S, Cornell B D, Greger B, Scherberger H and Andersen R A 2004 Cognitive control signals for neural prosthetics *Science* **305** 258–62
- Neumann N, Hinterberger T, Kaiser J, Leins U, Birbaumer N and Kubler A 2004 Automatic processing of self-regulation of slow cortical potentials: evidence from brain–computer communication in paralysed patients *Electroencephalogr. Clin. Neurophysiol.* **115** 628–35
- Neumann N, Kubler A, Kaiser J, Hinterberger T and Birbaumer N 2003 Conscious perception of brain states: Mental strategies for brain–computer communication *Neuropsychologia* **41** 1028–36
- Neuper C, Muller G R, Kubler A, Birbaumer N and Pfurtscheller G 2003 Clinical application of an EEG-based brain–computer interface: a case study in a patient with severe motor impairment *Clin. Neurophysiol.* **114** 399–409
- Neuper C, Scherer R, Reiner M and Pfurtscheller G 2005 Imagery of motor actions: differential effects of kinesthetic and visual-motor mode of imagery in single-trial EEG *Brain Res. Cogn. Brain Res.* **25** 668–77
- Neuper C, Schloegl A and Pfurtscheller G 1999 Enhancement of left-right sensorimotor EEG differences during feedback-regulated motor imagery *J. Clin. Neurophysiol.* **16** 373–82
- Niedermeyer E and Lopes da Silva F 1998 *Electroencephalography: Basic Principles, Clinical Applications, and Related Fields* 4th edn, ed C Mitchell (Baltimore, MD: Williams & Wilkins)
- Obeid I and Wolf P D 2004 Evaluation of spike-detection algorithms for a brain-machine interface application *IEEE Trans. Biomed. Eng.* **51** 905–11
- Obermaier B, Guger C, Neuper C and Pfurtscheller G 2001a Hidden markov models for online classification of single trial EEG data *Pattern Recognit. Lett.* **22** 1299–309
- Obermaier B, Muller G R and Pfurtscheller G 2001b ‘Virtual keyboard’ controlled by spontaneous EEG activity *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 636–41
- Obermaier B, Munteanu C, Rosa A and Pfurtscheller G 2001c Asymmetric hemisphere modeling in an offline brain–computer interface *IEEE Trans. Syst. Man Cybern.* **31** 536–40
- Obermaier B, Neuper C, Guger C and Pfurtscheller G 2001d Information transfer rate in a five-classes brain–computer interface *IEEE Trans. Neural Syst. Rehabil. Eng.* **9** 283–8
- Olson B P, Si J, Hu J and He J 2005 Closed-loop cortical control of direction using support vector machines *IEEE Trans. Neural Syst. Rehabil. Eng.* **13** 72–80
- Otto K J, Vetter R J, Marzullo T C and Kipke D R 2003 Brain-machine interfaces in rat motor cortex: implications of adaptive decoding algorithms *Proc. IEEE-EMBS Conf. on Neural Engineering (Capri Island, Italy)* pp 100–3
- Palaniappan R 2005 Brain–computer interface design using band powers extracted during mental tasks *Proc. IEEE-EMBS Conf. on Neural Engineering (Arlington, VA)* pp 321–4
- Palaniappan R, Paramesran R, Nishida S and Saiwaki N 2002 A new brain–computer interface design using fuzzy ARTMAP *IEEE Trans. Neural Syst. Rehabil. Eng.* **10** 140–8
- Panuccio A, Bicego M and Murino V 2002 A hidden markov model-based approach to sequential data clustering *Proc. Joint IAPR Int. Workshop on Structural, Syntactic, and Statistical Pattern Recognition (Windsor, Canada)* pp 734–42
- Parra L, Alvino C, Tang A, Pearlmutter B, Yeung N, Osman A and Sajda P 2002 Linear spatial integration for single-trial detection in encephalography *Neuroimage* **17** 223–30

- Parra L, Alvino C, Tang A, Pearlmutter B, Yeung N, Osman A and Sajda P 2003a Single-trial detection in EEG and MEG: keeping it linear *Neurocomputing* **52–54** 177–83
- Parra L C, Spence C D, Gerson A D and Sajda P 2003b Response error correction—a demonstration of improved human-machine performance using real-time EEG monitoring *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 173–7
- Patil P G, Carmena J M, Nicolelis M A and Turner D A 2004 Ensemble recordings of human subcortical neurons as a source of motor control signals for a brain-machine interface *Neurosurgery* **55** 27–35
- Penny W D, Roberts S J, Curran E A and Stokes M J 2000 EEG-based communication: a pattern recognition approach *IEEE Trans. Rehabil. Eng.* **8** 214–5
- Peters B O, Pfurtscheller G and Flyvbjerg H 2001 Automatic differentiation of multichannel EEG signals *IEEE Trans. Biomed. Eng.* **48** 111–6
- Peterson D A, Knight J N, Kirby M J, Anderson C W and Thaut M H 2005 Feature selection and blind source separation in an EEG-based brain–computer interface *EURASIP J. Appl. Signal Process.* **19** 3128–40
- Pfurtscheller G and Aranibar A 1977 Event-related cortical desynchronization detected by power measurements of scalp EEG *Electroencephalogr. Clin. Neurophysiol.* **42** 817–26
- Pfurtscheller G *et al* 2003b Graz-BCI: state of the art and clinical applications *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 177–80
- Pfurtscheller G, Flotzinger D and Kalcher J 1993 Brain–computer interface—a new communication device for handicapped persons *J. Microcomput. Appl.* **16** 293–9
- Pfurtscheller G, Flotzinger D and Neuper C 1994 Differentiation between finger, toe and tongue movement in man based on 40-Hz EEG *Electroencephalogr. Clin. Neurophysiol.* **90** 456–60
- Pfurtscheller G and Guger C 1999 Brain–computer communication system: EEG-based control of hand orthosis in a tetraplegic patient *Acta Chir. Austriaca* **31** 23–5
- Pfurtscheller G, Kalcher J, Neuper C, Flotzinger D and Pregenzer M 1996 On-line EEG classification during externally-paced hand movements using a neural network-based classifier *Electroencephalogr. Clin. Neurophysiol.* **99** 416–25
- Pfurtscheller G and Lopes da Silva F H 1999 Event-related EEG/MEG synchronization and desynchronization: basic principles *Clin. Neurophysiol.* **110** 1842–57
- Pfurtscheller G, Muller G R, Pfurtscheller J, Gerner H J and Rupp R 2003a ‘Thought’—control of functional electrical stimulation to restore hand grasp in a patient with tetraplegia *Neurosci. Lett.* **351** 33–6
- Pfurtscheller G, Muller-Putz G R, Pfurtscheller J and Rupp R 2005 EEG-based asynchronous BCI controls functional electrical stimulation in a tetraplegic patient *EURASIP J. Appl. Signal Process.* **19** 3152–5
- Pfurtscheller G and Neuper C 2001 Motor imagery and direct brain–computer communication *Proc. IEEE* **89** 1123–34
- Pfurtscheller G, Neuper C, Flotzinger D and Pregenzer M 1997 EEG-based discrimination between imagination of right and left hand movement *Electroencephalogr. Clin. Neurophysiol.* **103** 642–51
- Pfurtscheller G, Neuper C, Guger C, Harkam W, Ramoser H, Schlogl A, Obermaier B and Pregenzer M 2000 Current trends in graz brain–computer interface (BCI) research *IEEE Trans. Rehabil. Eng.* **8** 216–9
- Pfurtscheller G, Neuper C, Schlogl A and Lugger K 1998 Separability of EEG signals recorded during right and left motor imagery using adaptive autoregressive parameters *IEEE Trans. Rehabil. Eng.* **6** 316–25
- Pineda J A, Allison B Z and Vankov A 2000 The effects of self-movement, observation, and imagination on mu rhythms and readiness potentials (RP’s): toward a brain–computer interface (BCI) *IEEE Trans. Rehabil. Eng.* **8** 219–22
- Pineda J A, Silverman D S, Vankov A and Hestenes J 2003 Learning to control brain rhythms: making a brain–computer interface possible *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 181–4
- Platt J 1991 A resource-allocating network for function interpolation *Neural Comput.* **3** 213–25
- Polak M and Kostov A 1997 Development of brain–computer interface: preliminary results *Proc. 19th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Chicago, IL)* pp 1543–6
- Polak M and Kostov A 1998 Feature extraction in development of brain–computer interface: a case study *Proc. 20th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Hong Kong)* pp 2058–61
- Polak M and Kostov A 1999 Training setup for control of neural prosthesis using brain–computer interface *Proc. 21st Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society & Annual Fall Meeting of the Biomedical Engineering Society (Atlanta, GA)* p 446
- Polikoff J B, Bunnell H T and Borkowski W 1995 Toward a P300-based computer interface *Proc. RESNA’95 Annual Conf. (Vancouver, Canada)* pp 178–80
- Pregenzer M and Pfurtscheller G 1995 Distinction sensitive learning vector quantization (DSLQ)—application as a classifier based feature selection method for a brain–computer interface *Proc. 4th Int. Conf. on Artificial Neural Networks (Cambridge, UK)* pp 433–6
- Pregenzer M and Pfurtscheller G 1999 Frequency component selection for an EEG-based brain to computer interface *IEEE Trans. Rehabil. Eng.* **7** 413–9
- Qin L, Deng J, Ding L and He B 2004a Motor imagery classification by means of source analysis methods *Proc. 26th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (San Francisco, CA)* pp 4356–8
- Qin L, Ding L and He B 2004b Motor imagery classification by means of source analysis for brain–computer interface applications *J. Neural Eng.* **1** 135–41
- Qin L and He B 2005 A wavelet-based time-frequency analysis approach for classification of motor imagery for brain–computer interface applications *J. Neural Eng.* **2** 65–72
- Qin L, Kamousi B, Liu Z M, Ding L and He B 2005 Classification of motor imagery tasks by means of time-frequency-spatial analysis for brain–computer interface applications *Proc. IEEE-EMBS Conf. on Neural Engineering (Arlington, VA)* pp 374–6
- Ramachandran V S and Hirstein W 1998 The perception of phantom limbs *Brain* **121** 1603–30
- Ramoser H, Muller-Gerking J and Pfurtscheller G 2000 Optimal spatial filtering of single trial EEG during imagined hand movement *IEEE Trans. Rehabil. Eng.* **8** 441–6
- Rao Y N, Kim S, Sanchez J C, Erdogmus D, Principe J C, Carmena J M, Lebedev M A and Nicolelis M A 2005 Learning mappings in brain machine interfaces with echo state networks *Proc. IEEE Int. Conf. on Acoustics, Speech and Signal Processing (Philadelphia, PA)* pp v/233–v/236
- Reger B D, Fleming K M, Sanguineti V, Alford S and Mussa-Ivaldi F A 2000a Connecting brains to robots: the development of a hybrid system for the study of learning in neural tissues *Proc. 7th Int. Conf. on Artificial Life (Portland, OR)* pp 263–72
- Reger B D, Fleming K M, Sanguineti V, Alford S and Mussa-Ivaldi F A 2000b Connecting brains to robots: an artificial body for studying the computational properties of neural tissues *Artif. Life* **6** 307–24
- Rezek I, Roberts S and Sykacek P 2003 Ensemble coupled hidden markov models for joint characterization of dynamic signals *Proc. 9th Int. Workshop on Artificial Intelligence and Statistics (Key West, FL)* ed C M Bishop and B J Frey

- Roberts S J and Penny W 2003 Real-time brain–computer interfacing: a preliminary study using bayesian learning *Med. Biol. Eng. Comput.* **38** 56–61
- Roberts S J, Penny W and Rezek I 1999 Temporal and spatial complexity measures for electroencephalogram based brain–computer interfacing *Med. Biol. Eng. Comput.* **37** 93–8
- Roushe P, Clement R, Visser C, Cruz J and Si J 2003 ‘Neurobotics’: an integration of neurons, rats and robots for advanced studies of brain–computer interface systems *Proc. IEEE-EMBS Conf. on Neural Engineering (Capri Island, Italy)* pp 462–5
- Sanchez J C, Carmena J M, Lebedev M A, Nicolelis M A, Harris J G and Principe J C 2004 Ascertaining the importance of neurons to develop better brain-machine interfaces *IEEE Trans. Biomed. Eng.* **51** 943–53
- Sanchez J C, Erdogmus D, Rao Y N, Principe J and Nicolelis M A 2002a A comparison between nonlinear mappings and linear state estimation to model the relation from motor cortical neuronal firing to hand movements *Proc. SAB Workshop on Motor Control in Humans and Robots (Edinburgh, Scotland)* pp 59–65
- Sanchez J C, Erdogmus D, Rao Y N, Principe J, Nicolelis M A and Wessberg J 2003 Learning the contributions of the motor, premotor, and posterior parietal cortices for hand trajectory reconstruction in a brain machine interface *Proc. IEEE-EMBS Conf. on Neural Engineering (Capri Island, Italy)* pp 59–62
- Sanchez J C, Kim S P, Erdogmus D, Rao Y N, Principe J, Wessing J and Nicolelis M A 2002b Input–output mapping performance of linear and nonlinear models for estimating hand trajectories from cortical neuronal firing patterns *Proc. IEEE Workshop on Neural Networks and Signal Processing (Martigny, Switzerland)* pp 139–48
- Schalk G, Wolpaw J R, McFarland D J and Pfurtscheller G 2000 EEG-based communication: presence of an error potential *Clin. Neurophysiol.* **111** 2138–44
- Scherer R, Muller G R, Neuper C, Graimann B and Pfurtscheller G 2004 An asynchronously controlled EEG-based virtual keyboard: improvement of the spelling rate *IEEE Trans. Biomed. Eng.* **51** 979–84
- Schloegl A, Lugger K and Pfurtscheller G 1997a Using adaptive autoregressive parameters for a brain–computer-interface experiment *Proc. 19th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Chicago, IL)* pp 1533–5
- Schloegl A, Neuper C and Pfurtscheller G 1997b Subject specific EEG patterns during motor imagery *Proc. 19th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Chicago, IL)* pp 1530–2
- Schlogl A, Keinrath C, Scherer R and Pfurtscheller G 2003 Information transfer of an EEG-based brain–computer interface *Proc. IEEE-EMBS Conf. on Neural Engineering (Capri Island, Italy)* pp 641–4
- Schroder M, Bogdan M, Rosenstiel W, Hinterberger T and Birbaumer N 2003 Automated EEG feature selection for brain–computer interfaces *Proc. IEEE-EMBS Conf. on Neural Engineering (Capri Island, Italy)* pp 626–9
- Schröder M, Lal T N, Hinterberger T, Bogdan M, Hill N J, Birbaumer N, Rosenstiel W and Schölkopf B 2005 Robust EEG channel selection across subjects for brain–computer interfaces *EURASIP J. Appl. Signal Process.* **19** 3103–12
- Schwartz A B, Taylor D M and Tillery S I 2001 Extraction algorithms for cortical control of arm prosthetics *Curr. Opin. Neurobiol.* **11** 701–7
- Serby H, Yom-Tov E and Inbar G F 2005 An improved P300-based brain–computer interface *IEEE Trans. Neural Syst. Rehabil. Eng.* **13** 89–98
- Serruya M, Hatsopoulos N, Fellows M, Paninski L and Donoghue J 2003 Robustness of neuroprosthetic decoding algorithms *Biol. Cybern.* **88** 219–28
- Serruya M D, Hatsopoulos N G, Paninski L, Fellows M R and Donoghue J P 2002 Instant neural control of a movement signal *Nature* **416** 141–2
- Sheikh H, McFarland D J, Sarnacki W A and Wolpaw J R 2003 Electroencephalographic (EEG)-based communication: EEG control versus system performance in humans *Neurosci. Lett.* **345** 89–92
- Stone M 1977 Asymptotics for and against cross-validation *Biometrika* **64** 29–35
- Su Ryu C, Song Y, Yoo D S, Choi S, Moon S S and Sohn J H 1999 EEG-based discrimination between yes and no *Proc. 21st Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society & Annual Fall Meeting of the Biomedical Engineering Society (Atlanta, GA)* p 444
- Sutter E E 1992 The brain response interface: communication through visually-induced electrical brain responses *J. Microcomput. Appl.* **15** 31–45
- Sykacek P, Roberts S J and Stokes M 2004 Adaptive BCI based on variational bayesian kalman filtering: an empirical evaluation *IEEE Trans. Biomed. Eng.* **51** 719–27
- Sykacek P, Roberts S, Stokes M, Curran E, Gibbs M and Pickup L 2003 Probabilistic methods in BCI research *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 192–5
- Tavakolian K, Nasrabadi A M and Rezaei S 2004 Selecting better EEG channels for classification of mental tasks *Proc. IEEE Int. Symp. on Circuits and Systems* 537–40
- Taylor D M, Tillery S I and Schwartz A B 2002 Direct cortical control of 3D neuroprosthetic devices *Science* **296** 1829–32
- Taylor D M, Tillery S I H and Schwartz A B 2003 Information conveyed through brain-control: cursor versus robot *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 195–9
- Thulasidas M, Guan C, Ranganatha S, Wu J K, Zhu X and Xu W 2004 Effect of ocular artifact removal in brain–computer interface accuracy *Proc. 26th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (San Francisco, CA)* pp 4385–8
- Townsend G, Graimann B and Pfurtscheller G 2004 Continuous EEG classification during motor imagery—simulation of an asynchronous BCI *IEEE Trans. Neural Syst. Rehabil. Eng.* **12** 258–65
- Trejo L J, Wheeler K R, Jorgensen C C, Rosipal R, Clanton S T, Matthews B, Hibbs A D, Matthews R and Krupka M 2003 Multimodal neuroelectric interface development *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 199–204
- Varsta M, Heikkonen J, Millan J D R and Mourino J 2000 Evaluating the performance of three feature sets for brain–computer interfaces with an early stopping MLP committee *Int. Conf. on Pattern Recognition (Barcelona, Spain)* pp 907–10
- Vetter R J, Otto K J, Marzullo T C and Kipke D R 2003 Brain machine interfaces in rat motor cortex: neuronal operant conditioning to perform a sensory detection task *Proc. IEEE-EMBS Conf. on Neural Engineering (Capri Island, Italy)* pp 637–40
- Vidal J J 1977 Real-time detection of brain events in EEG *Proc. IEEE* **65** 633–41
- Wang T, Deng J and He B 2004a Classifying EEG-based motor imagery tasks by means of time-frequency synthesized spatial patterns *Clin. Neurophysiol.* **115** 2744–53
- Wang T, Deng J and He B 2004b Classification of motor imagery EEG patterns and their topographic representation *Proc. 26th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (San Francisco, CA)* pp 4359–62
- Wang D, Kochiyama T, Lu S and Wu J 2005a Measurement and analysis of electroencephalogram (EEG) using directional visual stimuli for brain–computer interface *Proc. Int. Conf. on Active Media Technology (Takamatsu, Japan)* pp 34–9
- Wang Y, Wang R, Gao X and Gao S 2005b Brain–computer interface based on the high-frequency steady-state visual

- evoked potential *Proc. 1st Int. Conf. on Neural Interface and Control (Wuhan, China)* pp 37–9
- Wang Y, Zhang Z, Gao X and Gao S 2004c Lead selection for SSVEP-based brain–computer interface *Proc. 26th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (San Francisco, CA)* pp 4507–10
- Wang Y, Zhang Z, Li Y, Gao X, Gao S and Yang F 2004d BCI competition 2003–data set IV: an algorithm based on CSSD and FDA for classifying single-trial EEG *IEEE Trans. Biomed. Eng.* **51** 1081–6
- Weiskopf N, Mathiak K, Bock S W, Scharnowski F, Veit R, Grodd W, Goebel R and Birbaumer N 2004 Principles of a brain–computer interface (BCI) based on real-time functional magnetic resonance imaging (fMRI) *IEEE Trans. Biomed. Eng.* **51** 966–70
- Weiskopf N, Veit R, Erb M, Mathiak K, Grodd W, Goebel R and Birbaumer N 2003 Physiological self-regulation of regional brain activity using real-time functional magnetic resonance imaging (fMRI): methodology and exemplary data *Neuroimage* **19** 577–86
- Weitkunat R 1991 *Digital Biosignal Processing* (New York: Elsevier)
- Wessberg J, Stambaugh C R, Kralik J D, Beck P D, Laubach M, Chapin J K, Kim J, Biggs S J, Srinivasan M A and Nicolelis M A 2000 Real-time prediction of hand trajectory by ensembles of cortical neurons in primates *Nature* **408** 361–5
- Wolpaw J R, Birbaumer N, McFarland D J, Pfurtscheller G and Vaughan T M 2002 Brain–computer interfaces for communication and control *Clin. Neurophysiol.* **113** 767–91
- Wolpaw J R, Flotzinger D, Pfurtscheller G and McFarland D J 1997 Timing of EEG-based cursor control *J. Clin. Neurophysiol.* **14** 529–38
- Wolpaw J R and McFarland D J 2004 Control of a two-dimensional movement signal by a noninvasive brain–computer interface in humans *Proc. Natl Acad. Sci. USA* **101** 17849–54
- Wolpaw J R and McFarland D J 1994 Multichannel EEG-based brain–computer communication *Electroencephalogr. Clin. Neurophysiol.* **90** 444–9
- Wolpaw J R, McFarland D, Neat G W and Forneris C A 1991 An EEG-based brain–computer interface for cursor control *Electroencephalogr. Clin. Neurophysiol.* **78** 252–9
- Wolpaw J R, McFarland D J and Vaughan T M 2000 Brain–computer interface research at the wadsworth center *IEEE Trans. Rehabil. Eng.* **8** 222–6
- Wolpaw J R, McFarland D J, Vaughan T M and Schalk G 2003 The wadsworth center brain–computer interface (BCI) research and development program *IEEE Trans. Neural Syst. Rehabil. Eng.* **11** 204–7
- Wu W, Black M J, Gao Y, Bienenstock E, Serruya M D and Donoghue J P 2002a Inferring hand motion from multi-cell recordings in motor cortex using a kalman filter *Proc. SAB Workshop on Motor Control in Humans and Robots (Edinburgh, Scotland)* pp 66–73
- Wu W, Black M J, Gao Y, Bienenstock E, Shaikhounix A and Donoghue J P 2002b Neural decoding of cursor motion using a kalman filter *Adv. Neural Inform. Process. Syst.* **15** 133–40
- Wu R C, Liang S F, Lin C T and Hsu C F 2004 Applications of event-related-potential-based brain–computer interface to intelligent transportation systems *Proc. IEEE Int. Conf. on Networking, Sensing and Control (Taipei, Taiwan)* pp 813–8
- Xu N, Gao X, Hong B, Miao X, Gao S and Yang F 2004a BCI competition 2003–data set IIb: enhancing P300 wave detection using ICA-based subspace projections for BCI applications *IEEE Trans. Biomed. Eng.* **51** 1067–72
- Xu W, Guan C, Siong C E, Ranganatha S, Thulasidas M and Wu J 2004b High accuracy classification of EEG signal *Proc. 17th Int. Conf. on Pattern Recognition (Cambridge, UK)* pp 391–4
- Yan W, Sutherland M T, Sanfratello L L and Tang A C 2004 Single-trial classification of ERPS using second-order blind identification (SOBI) *Proc. IEEE Int. Conf. on Machine Learning and Cybernetics (Shanghai, China)* pp 4246–51
- Yom-Tov E and Inbar G F 2001 Selection of relevant features for classification of movements from single movement-related potentials using a genetic algorithm *Proc. 23rd Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Istanbul, Turkey)* pp 1364–6
- Yom-Tov E and Inbar G F 2002 Feature selection for the classification of movements from single movement-related potentials *IEEE Trans. Neural Syst. Rehabil. Eng.* **10** 170–7
- Yom-Tov E and Inbar G F 2003 Detection of movement-related potentials from the electro-encephalogram for possible use in a brain–computer interface *Med. Biol. Eng. Comput.* **41** 85–93
- Yoo S S, Fairney T, Chen N K, Choo S E, Panych L P, Park H, Lee S Y and Jolesz F A 2004 Brain–computer interface using fMRI: spatial navigation by thoughts *Neuroreport* **15** 1591–5
- Yoon H, Yang K and Shahabi C 2005 Feature subset selection and feature ranking for multivariate time series *IEEE Trans. Knowl. Data Eng.* **17** 1186–98
- Yu Z, Mason S G and Birch G E 2002 Enhancing the performance of the LF-ASD brain–computer interface *Proc. 24th Annual Int. Conf. of the IEEE Engineering in Medicine and Biology Society (Houston, TX)* pp 2443–4